



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101524\
Data File : VW030598.D
Acq On : 15 Oct 2024 08:23
Operator : SY/MD
Sample : BFB576
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB576

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Title : SFAM01.0
Last Update : Wed Oct 16 01:28:14 2024



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1752

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	23598	PASS
75	95	30	80	58.1	58751	PASS
95	95	100	100	100.0	101144	PASS
96	95	5	9	7.1	7182	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	76.0	76883	PASS
175	174	5	9	8.5	6505	PASS
176	174	95	101	97.0	74563	PASS
177	176	5	9	6.4	4784	PASS

m/z	Abundance
35.40	61.0
36.10	84.0
36.90	76.0
37.90	88.0
38.70	60.0
40.10	273.0
41.90	87.0
43.00	71.0
44.00	192.0
47.00	84.0
48.20	60.0
50.10	61.0
53.20	53.0
57.10	59.0
69.90	53.0
74.90	70.0
88.20	53.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

m/z	Abundance
36.00	1362.0
37.10	6405.0
38.00	5760.0
39.10	2492.0
40.10	386.0
41.40	98.0
43.00	56.0
44.00	1148.0
45.10	1166.0
45.90	199.0
47.00	1101.0
48.00	910.0
49.10	5569.0
50.10	23320.0
51.10	6818.0
52.10	387.0
53.10	185.0
53.90	63.0
55.10	366.0
56.10	2081.0
57.10	3530.0
58.30	126.0
58.90	157.0
60.00	1238.0
61.10	6055.0
62.00	5995.0
63.10	4982.0
64.00	892.0
65.10	978.0
66.00	113.0
67.10	410.0
68.10	11714.0
69.10	12138.0
70.10	1000.0
72.10	789.0
73.00	5505.0
74.10	20264.0
75.10	53280.0
76.10	5286.0
77.00	519.0
78.00	418.0
79.00	5083.0
80.00	1317.0
81.00	4985.0
82.00	1212.0
83.00	255.0
85.90	184.0
87.00	4364.0
88.00	3911.0
90.90	498.0
92.00	3167.0
93.00	4894.0
94.10	12368.0
95.10	91328.0
96.10	7233.0
97.00	229.0
102.90	121.0
104.00	582.0
105.10	180.0
106.00	502.0
107.00	132.0
110.00	92.0
110.80	159.0
111.50	63.0
113.00	121.0
114.90	120.0
116.00	435.0
117.00	853.0
118.00	428.0
118.90	685.0
121.90	80.0
123.80	92.0
124.80	193.0
125.20	161.0
125.70	80.0
128.00	391.0
128.80	215.0
130.10	305.0
130.90	171.0
132.00	68.0
132.90	435.0
135.10	202.0
135.90	65.0
136.90	185.0
140.10	142.0
141.00	1080.0
142.00	161.0
143.00	1497.0
144.00	137.0
145.00	406.0
145.70	204.0
147.00	137.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

148.00	204.0
148.90	100.0
150.00	145.0
151.90	101.0
152.80	66.0
153.10	68.0
153.90	89.0
154.80	234.0
156.00	65.0
156.90	149.0
159.10	190.0
160.90	144.0
162.80	88.0
169.50	111.0
170.30	75.0
171.00	279.0
171.90	1147.0
174.00	60456.0
175.10	5581.0
176.00	61584.0
177.00	4054.0
178.00	210.0
178.90	110.0
191.10	335.0
192.30	119.0
193.00	505.0
194.00	106.0
194.90	68.0
203.20	66.0
205.10	61.0
207.10	339.0
234.60	56.0
235.20	69.0
249.00	228.0
250.80	93.0
265.10	339.0
265.90	50.0
279.60	109.0
281.20	3260.0
282.20	736.0
283.00	630.0
285.10	88.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

m/z	Abundance
36.00	1407.0
37.10	7494.0
38.10	6441.0
39.10	2597.0
40.00	448.0
40.80	112.0
41.40	76.0
42.00	67.0
42.40	54.0
43.00	126.0
44.00	1302.0
45.00	1230.0
46.00	249.0
47.00	1365.0
48.00	971.0
49.10	6177.0
50.10	26728.0
51.10	7886.0
52.00	451.0
53.20	124.0
55.10	389.0
56.10	2371.0
57.00	4306.0
57.90	231.0
59.00	115.0
60.10	1286.0
61.00	7347.0
62.10	7120.0
63.10	5597.0
64.00	906.0
65.10	733.0
65.80	147.0
67.00	369.0
68.10	13967.0
69.00	14638.0
70.10	986.0
71.10	92.0
72.00	949.0
73.00	6839.0
74.10	23496.0
75.10	66584.0
76.10	5556.0
77.00	665.0
78.00	604.0
79.00	5526.0
80.00	1604.0
81.00	6044.0
82.00	1280.0
82.90	134.0
86.10	236.0
87.00	5359.0
88.00	5062.0
91.00	577.0
92.00	4068.0
93.10	5907.0
94.10	15546.0
95.10	114680.0
96.10	7363.0
97.10	238.0
102.90	219.0
103.90	666.0
104.90	151.0
106.00	623.0
106.90	171.0
110.00	150.0
111.10	216.0
112.10	124.0
112.90	83.0
114.80	157.0
116.00	387.0
117.00	1151.0
118.00	657.0
119.00	995.0
121.90	70.0
123.90	80.0
125.00	214.0
125.90	81.0
127.10	68.0
127.90	434.0
128.90	236.0
130.00	476.0
131.10	256.0
132.80	214.0
133.30	211.0
134.00	103.0
135.00	318.0
135.90	63.0
136.90	284.0
138.80	117.0
139.90	116.0
141.00	1332.0
142.00	239.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

143.00	1691.0
144.20	99.0
144.90	359.0
145.90	214.0
147.10	190.0
148.00	316.0
148.80	64.0
149.70	71.0
149.90	71.0
151.90	65.0
152.80	118.0
154.00	94.0
154.90	267.0
156.00	65.0
156.80	265.0
158.80	173.0
160.80	106.0
161.10	101.0
162.80	62.0
165.00	70.0
168.80	89.0
170.00	146.0
170.40	121.0
171.10	172.0
172.00	1287.0
174.00	83808.0
175.00	7897.0
176.00	78280.0
177.00	5042.0
178.00	168.0
179.00	97.0
190.70	120.0
191.00	109.0
191.80	58.0
193.00	393.0
194.00	89.0
207.00	284.0
249.00	171.0
250.80	94.0
265.00	292.0
266.90	97.0
279.50	69.0
280.20	99.0
281.20	2318.0
282.10	532.0
283.10	500.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

m/z	Abundance
36.00	1192.0
37.10	6240.0
38.10	4914.0
39.10	1843.0
40.10	476.0
40.90	75.0
41.90	74.0
43.10	157.0
44.10	985.0
45.10	1015.0
45.90	265.0
47.00	1143.0
48.00	838.0
49.10	4638.0
50.10	20928.0
51.10	6554.0
52.00	318.0
53.00	85.0
55.00	285.0
56.10	2094.0
57.10	3779.0
58.10	155.0
58.70	117.0
60.00	1374.0
61.10	5866.0
62.10	6022.0
63.10	4490.0
64.10	490.0
65.00	784.0
66.00	103.0
67.00	412.0
68.00	10857.0
69.10	11968.0
70.10	951.0
72.10	673.0
73.00	5138.0
74.10	19376.0
75.10	56600.0
76.10	5410.0
77.20	606.0
78.00	596.0
79.00	5066.0
80.00	1322.0
81.00	4523.0
82.00	1125.0
83.10	140.0
84.30	56.0
85.70	59.0
86.10	115.0
87.00	4305.0
87.90	4513.0
91.00	483.0
92.10	3147.0
93.10	5620.0
94.10	13699.0
95.10	97424.0
96.10	6949.0
97.00	176.0
104.00	583.0
105.00	252.0
105.90	518.0
106.90	144.0
110.00	139.0
111.00	181.0
112.10	122.0
112.90	187.0
114.60	87.0
116.00	520.0
116.90	935.0
118.00	462.0
118.90	750.0
119.80	66.0
121.90	66.0
124.00	93.0
125.00	58.0
126.00	64.0
126.50	57.0
128.00	485.0
128.90	125.0
129.90	366.0
130.90	264.0
133.10	124.0
133.80	81.0
134.20	105.0
134.90	252.0
136.90	223.0
139.00	111.0
139.90	127.0
141.00	1330.0
141.90	220.0
143.00	1283.0
144.20	77.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576

144.90	260.0
146.00	255.0
147.00	119.0
148.00	292.0
148.90	71.0
149.90	144.0
152.00	80.0
152.90	138.0
155.00	190.0
156.90	326.0
159.00	161.0
161.00	160.0
165.00	53.0
168.80	77.0
170.40	65.0
171.00	248.0
172.00	1041.0
174.00	86384.0
175.00	6038.0
176.00	83824.0
177.00	5256.0
178.10	207.0
191.00	92.0
192.90	201.0
194.10	82.0
207.00	246.0
249.00	72.0
264.80	105.0
265.20	99.0
279.40	63.0
281.10	1159.0
282.00	376.0
283.00	193.0
284.00	59.0

Instrument :
MSVOA_W
ClientSampleId :
BFB576