



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101624\
Data File : VW030611.D
Acq On : 16 Oct 2024 14:20
Operator : SY/MD
Sample : BFB577
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB577

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Title : SFAM01.0
Last Update : Thu Oct 17 01:53:12 2024



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1752

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.6	17363	PASS
75	95	30	80	56.9	43712	PASS
95	95	100	100	100.0	76839	PASS
96	95	5	9	6.5	4980	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.5	56453	PASS
175	174	5	9	8.4	4715	PASS
176	174	95	101	97.1	54795	PASS
177	176	5	9	6.7	3649	PASS

m/z	Abundance
36.40	55.0
37.40	68.0
37.90	74.0
38.50	73.0
39.40	71.0
40.00	302.0
40.80	63.0
41.50	112.0
42.20	62.0
43.20	57.0
44.10	288.0
44.90	54.0
49.60	61.0
51.40	63.0
52.80	58.0
54.60	71.0
56.80	70.0
61.00	52.0
64.10	57.0
65.30	61.0
71.30	52.0
95.20	62.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

m/z	Abundance
36.00	940.0
37.10	4805.0
38.00	4573.0
39.10	2093.0
40.00	344.0
40.80	94.0
41.40	58.0
42.20	65.0
43.20	135.0
44.10	843.0
45.10	931.0
45.80	131.0
46.40	153.0
47.00	1088.0
48.00	864.0
49.10	3957.0
50.10	17984.0
51.10	5410.0
52.00	372.0
52.90	191.0
55.00	298.0
56.00	1507.0
57.10	2885.0
58.10	217.0
58.90	109.0
60.00	876.0
61.10	4904.0
62.10	4873.0
63.10	3723.0
63.90	824.0
65.00	1081.0
66.20	183.0
67.00	320.0
68.10	7936.0
69.10	9543.0
70.10	778.0
71.30	141.0
72.10	515.0
73.10	4331.0
74.10	14891.0
75.10	41360.0
76.10	3808.0
77.00	529.0
78.00	471.0
79.00	4540.0
80.00	1041.0
81.00	4227.0
82.00	934.0
82.90	128.0
85.20	51.0
86.00	189.0
87.00	2991.0
88.00	2919.0
91.00	444.0
92.10	2384.0
93.00	3894.0
94.10	10072.0
95.00	76080.0
96.00	4928.0
97.00	227.0
103.00	75.0
104.00	361.0
105.20	107.0
105.90	401.0
108.10	66.0
109.90	82.0
110.30	95.0
111.00	101.0
112.90	153.0
115.00	122.0
116.00	243.0
117.00	744.0
118.00	404.0
118.90	512.0
125.10	86.0
127.90	265.0
129.00	133.0
130.00	283.0
130.80	118.0
133.00	310.0
135.00	160.0
136.90	106.0
141.00	695.0
141.60	158.0
142.20	165.0
142.90	1351.0
143.90	163.0
144.90	557.0
145.90	163.0
146.90	103.0
147.80	139.0
148.90	72.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

150.00	83.0
153.10	85.0
155.10	173.0
157.00	152.0
159.00	100.0
161.00	85.0
163.10	79.0
165.00	99.0
170.00	52.0
171.00	348.0
171.90	1193.0
174.00	47400.0
175.00	3789.0
176.00	50200.0
177.00	2969.0
178.00	157.0
178.90	73.0
191.10	251.0
192.80	314.0
199.80	61.0
207.20	238.0
249.10	186.0
250.90	130.0
265.10	331.0
266.20	51.0
267.10	154.0
279.00	81.0
281.00	2663.0
282.10	500.0
283.10	375.0
284.30	66.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

m/z	Abundance
35.20	53.0
36.00	1188.0
37.00	5724.0
38.10	5113.0
39.10	2409.0
39.90	313.0
41.00	71.0
42.10	85.0
43.10	98.0
44.00	1095.0
45.10	878.0
46.20	170.0
47.20	940.0
47.90	710.0
49.10	4864.0
50.10	18752.0
51.10	6012.0
52.10	385.0
53.10	208.0
54.00	115.0
55.00	286.0
56.10	1878.0
57.00	3284.0
58.00	137.0
58.50	104.0
60.10	1006.0
61.10	5323.0
62.10	5347.0
63.10	4490.0
64.10	642.0
65.10	1355.0
66.10	128.0
66.80	289.0
67.20	257.0
68.00	10222.0
69.10	10461.0
70.00	653.0
71.20	128.0
72.10	572.0
73.10	5439.0
74.10	16209.0
75.10	50232.0
76.10	4136.0
77.10	754.0
78.10	545.0
79.00	4772.0
80.00	1445.0
81.00	4589.0
82.00	936.0
83.20	116.0
85.80	224.0
87.00	3993.0
87.90	3533.0
89.90	70.0
91.00	598.0
92.10	2954.0
93.00	4409.0
94.10	12061.0
95.10	82520.0
96.00	5409.0
97.30	111.0
103.20	127.0
103.90	527.0
105.10	116.0
105.90	607.0
107.10	51.0
108.00	84.0
108.70	57.0
109.90	84.0
111.10	150.0
111.90	80.0
112.90	158.0
115.00	148.0
115.90	439.0
117.00	892.0
118.00	372.0
119.00	788.0
123.20	64.0
124.10	70.0
125.00	74.0
127.90	258.0
128.90	230.0
130.00	357.0
131.00	143.0
132.90	108.0
133.40	119.0
135.00	220.0
136.90	228.0
139.10	97.0
141.00	1118.0
142.00	189.0
143.00	1412.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

144.10	286.0
145.00	573.0
146.00	156.0
147.00	128.0
147.90	148.0
149.00	133.0
150.00	102.0
151.70	56.0
153.10	173.0
153.80	64.0
154.90	190.0
157.00	101.0
158.70	116.0
161.10	107.0
165.00	94.0
170.60	138.0
171.00	270.0
172.00	1109.0
174.00	65344.0
175.00	5543.0
176.00	55576.0
177.00	4203.0
177.90	83.0
191.10	108.0
193.10	238.0
205.50	62.0
207.20	106.0
249.10	105.0
250.90	91.0
264.80	124.0
281.10	1231.0
282.20	394.0
283.10	210.0
284.20	68.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

m/z	Abundance
35.20	64.0
36.10	900.0
37.10	3851.0
38.10	3883.0
39.00	1620.0
40.00	255.0
40.80	124.0
42.00	87.0
43.00	98.0
44.00	792.0
45.10	777.0
46.10	243.0
47.00	630.0
48.00	544.0
49.10	3401.0
50.10	15352.0
51.10	4975.0
52.00	273.0
53.10	244.0
53.90	96.0
55.10	236.0
56.10	1586.0
57.10	2203.0
58.10	148.0
59.10	60.0
60.00	1020.0
61.10	4443.0
62.10	4300.0
63.10	3845.0
64.10	573.0
65.00	795.0
66.10	127.0
67.00	208.0
68.10	8168.0
69.10	8325.0
70.10	682.0
72.00	579.0
73.10	3808.0
74.10	14853.0
75.10	39544.0
76.10	2896.0
77.10	532.0
79.00	3359.0
80.00	950.0
81.00	3546.0
82.00	922.0
83.20	66.0
86.20	133.0
87.00	2894.0
88.00	2942.0
90.00	51.0
91.10	308.0
92.00	2063.0
93.00	3866.0
94.10	10334.0
95.10	72104.0
96.10	4603.0
96.90	184.0
103.20	132.0
104.00	466.0
104.90	168.0
106.00	404.0
106.60	113.0
107.10	59.0
107.90	83.0
111.10	74.0
111.90	89.0
112.90	116.0
114.80	80.0
116.00	419.0
117.00	784.0
117.90	224.0
118.90	428.0
123.10	61.0
123.90	113.0
125.00	60.0
127.90	304.0
128.80	181.0
130.00	348.0
131.00	129.0
132.70	103.0
133.60	62.0
135.00	205.0
135.60	82.0
136.90	247.0
141.00	804.0
142.10	128.0
143.00	963.0
143.90	54.0
144.10	56.0
145.00	457.0
146.00	107.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577

147.00	112.0
147.90	157.0
149.10	90.0
150.10	83.0
152.50	62.0
153.10	58.0
153.80	107.0
154.90	182.0
156.90	176.0
158.90	157.0
161.00	196.0
170.20	68.0
171.10	190.0
172.00	1378.0
174.00	56616.0
175.00	4812.0
176.00	58608.0
177.00	3775.0
178.10	138.0
190.90	69.0
193.10	86.0
205.50	57.0
207.20	62.0
281.10	402.0
282.30	118.0
283.10	91.0
284.10	56.0

Instrument :
MSVOA_W
ClientSampleId :
BFB577