

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101823\
 Data File : VW027324.D
 Acq On : 18 Oct 2023 12:49
 Operator : SY/MD
 Sample : VIBLK533
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK533

Quant Time: Oct 18 22:22:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 18 22:20:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	438386	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	398553	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	198397	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	158613	23.707	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.840%
7) Chloroethane-d5	2.899	69	121785	25.351	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.400%
11) 1,1-Dichloroethene-d2	4.021	65	80710	24.075	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	96.320%
21) 2-Butanone-d5	7.075	46	87086	39.162	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.320%
24) Chloroform-d	7.648	84	313226	23.973	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.880%
26) 1,2-Dichloroethane-d4	8.306	65	163344	23.272	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.080%
32) Benzene-d6	8.276	84	668181	24.780	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.120%
36) 1,2-Dichloropropane-d6	9.276	67	201821	24.504	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.000%
41) Toluene-d8	10.324	98	597583	24.439	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.760%
43) trans-1,3-Dichloroprop...	10.574	79	80275	21.377	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.520%
47) 2-Hexanone-d5	10.922	63	70971	41.649	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	135793	22.099	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.400%
66) 1,2-Dichlorobenzene-d4	13.848	152	164687	22.906	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.640%
Target Compounds						Qvalue
16) Methylene chloride	4.929	84	15160	1.889	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101823\
Data File : VW027324.D
Acq On : 18 Oct 2023 12:49
Operator : SY/MD
Sample : VIBLK533
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK533

Quant Time: Oct 18 22:22:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 18 22:20:28 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 1.889 ug/L
RT: 4.929 min Scan# 49
Delta R.T. 0.006 min
Lab File: VW027324.D
Acq: 18 Oct 2023 12:49

Instrument :
MSVOA_W
ClientSampleId :
VIBLK533

Tgt Ion: 84 Resp: 15160
Ion Ratio Lower Upper
84 100
86 55.6 41.3 76.7
49 120.8 92.8 172.4

