

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101824\
 Data File : VW030663.D
 Acq On : 18 Oct 2024 14:41
 Operator : SY/MD
 Sample : P4429-15
 Misc : 4.93g/13mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 19 00:27:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 19 00:23:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	484594	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	395391	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	148924	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	84878	11.143	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	44.560%
7) Chloroethane-d5	2.899	69	88693	15.383	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.520%
11) 1,1-Dichloroethene-d2	4.021	65	33765	10.658	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	42.640%#
21) 2-Butanone-d5	7.075	46	46272	33.756	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	67.520%
24) Chloroform-d	7.648	84	249575	18.009	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.040%
26) 1,2-Dichloroethane-d4	8.307	65	124022	17.210	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	68.840%#
32) Benzene-d6	8.276	84	381675	16.632	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	66.520%
36) 1,2-Dichloropropane-d6	9.276	67	136212	20.223	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.880%
41) Toluene-d8	10.325	98	299333	14.288	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	57.160%
43) trans-1,3-Dichloroprop...	10.575	79	46048	16.327	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.320%
47) 2-Hexanone-d5	10.922	63	35700	41.151	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	99559	20.127	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.520%
66) 1,2-Dichlorobenzene-d4	13.855	152	93706	18.596	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101824\
Data File : VW030663.D
Acq On : 18 Oct 2024 14:41
Operator : SY/MD
Sample : P4429-15
Misc : 4.93g/13mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Oct 19 00:27:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 19 00:23:37 2024
Response via : Initial Calibration

