

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
 Data File : VW024950.D
 Acq On : 19 Oct 2022 20:22
 Operator : SY/VA
 Sample : N5193-01
 Misc : 3.62g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 LSP-SS-01-00-20

Quant Time: Oct 20 02:03:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 20 01:56:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	520672	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	442756	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	154806	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	106965	17.413	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.640%
7) Chloroethane-d5	2.886	69	70668	19.649	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	4.019	63	181471	13.587	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.360%
21) 2-Butanone-d5	7.080	46	115483	59.334	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.660%
24) Chloroform-d	7.647	84	345805	23.547	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.200%
26) 1,2-Dichloroethane-d4	8.305	65	182017	23.682	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.720%
32) Benzene-d6	8.275	84	662696	23.869	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.480%
36) 1,2-Dichloropropane-d6	9.275	67	223304	26.223	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.880%
41) Toluene-d8	10.323	98	539033	21.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.520%
43) trans-1,3-Dichloroprop...	10.579	79	73950	22.897	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.600%
47) 2-Hexanone-d5	10.921	63	87832	61.519	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.040%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	181862	27.385	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.520%
66) 1,2-Dichlorobenzene-d4	13.847	152	140819	24.892	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
Data File : VW024950.D
Acq On : 19 Oct 2022 20:22
Operator : SY/VA
Sample : N5193-01
Misc : 3.62g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
LSP-SS-01-00-20

Quant Time: Oct 20 02:03:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 20 01:56:31 2022
Response via : Initial Calibration

