

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
 Data File : VW024941.D
 Acq On : 19 Oct 2022 16:44
 Operator : SY/VA
 Sample : N4757-03
 Misc : 4.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBWK0

Quant Time: Oct 20 02:00:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 20 01:56:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	492327	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	416162	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	158422	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	98410	16.943	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.760%
7) Chloroethane-d5	2.885	69	64715	19.030	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.120%
11) 1,1-Dichloroethene-d2	4.019	63	167908	13.296	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.200%
21) 2-Butanone-d5	7.080	46	111622	60.652	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.300%
24) Chloroform-d	7.647	84	317542	22.867	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.480%
26) 1,2-Dichloroethane-d4	8.305	65	174205	23.970	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.880%
32) Benzene-d6	8.275	84	606726	23.249	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	9.274	67	201772	25.208	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.840%
41) Toluene-d8	10.323	98	501046	21.391	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.560%
43) trans-1,3-Dichloroprop...	10.579	79	72796	23.980	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.920%
47) 2-Hexanone-d5	10.921	63	83969	62.572	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.140%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	167917	26.901	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	126116	21.784	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.120%
Target Compounds					Qvalue	
13) Acetone	4.135	43	9359	7.267	ug/L	77
16) Methylene chloride	4.922	84	33568	3.862	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
Data File : VW024941.D
Acq On : 19 Oct 2022 16:44
Operator : SY/VA
Sample : N4757-03
Misc : 4.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBWK0

Quant Time: Oct 20 02:00:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 20 01:56:31 2022
Response via : Initial Calibration

