

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
 Data File : VW024937.D
 Acq On : 19 Oct 2022 14:36
 Operator : SY/VA
 Sample : VW1019SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK429

Quant Time: Oct 20 01:59:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 20 01:56:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	427587	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	382028	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	184899	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	113318	22.464	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.840%
7) Chloroethane-d5	2.885	69	66346	22.463	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.840%
11) 1,1-Dichloroethene-d2	4.019	63	183940	16.770	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.080%
21) 2-Butanone-d5	7.079	46	78693	49.233	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.460%
24) Chloroform-d	7.653	84	292027	24.214	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.840%
26) 1,2-Dichloroethane-d4	8.305	65	153519	24.322	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.280%
32) Benzene-d6	8.274	84	596188	24.887	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.560%
36) 1,2-Dichloropropane-d6	9.274	67	185479	25.243	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.960%
41) Toluene-d8	10.323	98	532493	24.765	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.080%
43) trans-1,3-Dichloroprop...	10.579	79	72289	25.941	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.760%
47) 2-Hexanone-d5	10.926	63	61664	50.057	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.120%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	142131	24.804	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	170546	25.240	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101922\
Data File : VW024937.D
Acq On : 19 Oct 2022 14:36
Operator : SY/VA
Sample : VW1019SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK429

Quant Time: Oct 20 01:59:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 20 01:56:31 2022
Response via : Initial Calibration

