

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101923\
 Data File : VW027342.D
 Acq On : 19 Oct 2023 12:50
 Operator : SY/MD
 Sample : 04922-08
 Misc : 6.25g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAS3

Quant Time: Oct 20 02:48:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 20 02:46:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	532968	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	489731	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	236546	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	117946	14.501	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	58.000%		
7) Chloroethane-d5	2.893	69	93564	16.020	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	64.080%		
11) 1,1-Dichloroethene-d2	4.027	65	61209	15.018	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.080%		
21) 2-Butanone-d5	7.075	46	136612	50.531	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.060%		
24) Chloroform-d	7.648	84	278575	17.537	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	70.160%		
26) 1,2-Dichloroethane-d4	8.307	65	169507	19.865	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	79.440%		
32) Benzene-d6	8.276	84	573168	17.299	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	69.200%		
36) 1,2-Dichloropropane-d6	9.276	67	188637	18.639	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	74.560%		
41) Toluene-d8	10.325	98	483488	16.091	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	64.360%		
43) trans-1,3-Dichloroprop...	10.575	79	80358	17.415	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	69.680%		
47) 2-Hexanone-d5	10.922	63	106204	50.722	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.440%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	174712	23.139	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	92.560%		
66) 1,2-Dichlorobenzene-d4	13.848	152	146259	17.062	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	68.240%#		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	25341	2.598	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101923\
Data File : VW027342.D
Acq On : 19 Oct 2023 12:50
Operator : SY/MD
Sample : 04922-08
Misc : 6.25g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAS3

Quant Time: Oct 20 02:48:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 20 02:46:04 2023
Response via : Initial Calibration

