

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102323\
 Data File : VW027398.D
 Acq On : 23 Oct 2023 16:26
 Operator : SY/MD
 Sample : 04961-06
 Misc : 5.74g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E0LJ2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2023
 Supervised By :Mahesh Dadoda 10/24/2023

Quant Time: Oct 24 00:51:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 21 04:25:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	510503	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	332449	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	81899	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	133858	17.181	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	68.720%		
7) Chloroethane-d5	2.899	69	109642	19.599	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	78.400%		
11) 1,1-Dichloroethene-d2	4.021	65	67350	17.252	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	69.000%		
21) 2-Butanone-d5	7.081	46	121316	46.848	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	93.700%		
24) Chloroform-d	7.648	84	316896	20.827	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	83.320%		
26) 1,2-Dichloroethane-d4	8.301	65	182306m	22.305	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	89.200%		
32) Benzene-d6	8.276	84	617483	27.453	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	109.800%		
36) 1,2-Dichloropropane-d6	9.270	67	203298	29.592	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	118.360%		
41) Toluene-d8	10.319	98	429544	21.060	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	84.240%		
43) trans-1,3-Dichloroprop...	10.575	79	71923	22.961	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	91.840%		
47) 2-Hexanone-d5	10.922	63	85604	60.226	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	120.460%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	137014	26.731	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	106.920%		
66) 1,2-Dichlorobenzene-d4	13.849	152	53615	18.065	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	72.280%#		
Target Compounds						
34) Trichloroethene	9.093	95	148928	25.950	ug/L	93
46) Tetrachloroethene	10.861	164	3308343	834.843	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102323\
Data File : VW027398.D
Acq On : 23 Oct 2023 16:26
Operator : SY/MD
Sample : 04961-06
Misc : 5.74g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E0LJ2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2023
Supervised By :Mahesh Dadoda 10/24/2023

Quant Time: Oct 24 00:51:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 21 04:25:20 2023
Response via : Initial Calibration

