

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102323\
 Data File : VW027399.D
 Acq On : 23 Oct 2023 16:50
 Operator : SY/MD
 Sample : VIBLK543
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK543

Quant Time: Oct 24 00:51:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 21 04:25:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	399891	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	360417	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	184669	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	131708	21.581	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.320%
7) Chloroethane-d5	2.899	69	105173	24.001	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.000%
11) 1,1-Dichloroethene-d2	4.021	65	69492	22.724	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	90.880%
21) 2-Butanone-d5	7.075	46	97130	47.883	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.760%
24) Chloroform-d	7.648	84	287273	24.103	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	8.307	65	157552	24.608	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.440%
32) Benzene-d6	8.276	84	602053	24.690	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.760%
36) 1,2-Dichloropropane-d6	9.276	67	188641	25.328	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.320%
41) Toluene-d8	10.319	98	517992	23.425	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.720%
43) trans-1,3-Dichloroprop...	10.575	79	71734	21.124	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.480%
47) 2-Hexanone-d5	10.922	63	73547	47.728	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.460%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	140811	25.340	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.360%
66) 1,2-Dichlorobenzene-d4	13.848	152	157086	23.473	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102323\
Data File : VW027399.D
Acq On : 23 Oct 2023 16:50
Operator : SY/MD
Sample : VIBLK543
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK543

Quant Time: Oct 24 00:51:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 21 04:25:20 2023
Response via : Initial Calibration

