

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102423\
 Data File : VW027408.D
 Acq On : 24 Oct 2023 11:43
 Operator : SY/MD
 Sample : 04962-11
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Oct 25 01:29:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 25 01:28:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	372611	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	339960	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	165368	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	124063	21.817	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	87.280%	
7) Chloroethane-d5	2.899	69	103330	25.306	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	101.240%	
11) 1,1-Dichloroethene-d2	4.027	65	64396	22.600	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	90.400%	
21) 2-Butanone-d5	7.075	46	92340	48.854	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	97.700%	
24) Chloroform-d	7.648	84	278186	25.049	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	100.200%	
26) 1,2-Dichloroethane-d4	8.307	65	152417	25.549	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	102.200%	
32) Benzene-d6	8.276	84	578738	25.162	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	100.640%	
36) 1,2-Dichloropropane-d6	9.270	67	184402	26.248	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	105.000%	
41) Toluene-d8	10.319	98	497524	23.853	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	95.400%	
43) trans-1,3-Dichloroprop...	10.575	79	66026	20.613	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	82.440%	
47) 2-Hexanone-d5	10.922	63	67612	46.517	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	93.040%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	134158	25.595	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	102.400%	
66) 1,2-Dichlorobenzene-d4	13.848	152	150150	25.056	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	100.240%	
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	15400	2.258	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102423\
Data File : VW027408.D
Acq On : 24 Oct 2023 11:43
Operator : SY/MD
Sample : 04962-11
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Oct 25 01:29:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 25 01:28:28 2023
Response via : Initial Calibration

