

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102423\
 Data File : VW027424.D
 Acq On : 24 Oct 2023 18:09
 Operator : SY/MD
 Sample : VIBLK548
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK548

Quant Time: Oct 25 01:33:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 25 01:28:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	382592	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	350639	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	174525	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	100942	17.288	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.160%
7) Chloroethane-d5	2.899	69	82065	19.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.280%
11) 1,1-Dichloroethene-d2	4.027	65	51983	17.767	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.080%
21) 2-Butanone-d5	7.081	46	88661	45.684	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.360%
24) Chloroform-d	7.648	84	232955	20.429	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.720%
26) 1,2-Dichloroethane-d4	8.307	65	130153	21.248	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.000%
32) Benzene-d6	8.276	84	474072	19.983	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.920%
36) 1,2-Dichloropropane-d6	9.276	67	153452	21.178	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.720%
41) Toluene-d8	10.319	98	406277	18.885	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.560%
43) trans-1,3-Dichloroprop...	10.575	79	56618	17.138	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.560%
47) 2-Hexanone-d5	10.922	63	66611	44.432	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.860%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	124986	23.119	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	133412	21.095	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.360%
Target Compounds						Qvalue
16) Methylene chloride	4.911	84	23445	3.348	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102423\
Data File : VW027424.D
Acq On : 24 Oct 2023 18:09
Operator : SY/MD
Sample : VIBLK548
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK548

Quant Time: Oct 25 01:33:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 25 01:28:28 2023
Response via : Initial Calibration

