

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102618\
 Data File : VW006363.D
 Acq On : 25 Oct 2018 14:39
 Operator : SY/AP
 Sample : J5657-03
 Misc : 5.96G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A40K1

Quant Time: Oct 26 04:47:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100918S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 26 04:05:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	370134	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	322651	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	147018	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	74567	17.86	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.44%
7) Chloroethane-d5	2.90	69	71777	19.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.92%
10) 1,1-Dichloroethene-d2	4.01	63	117773	14.69	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.76%
20) 2-Butanone-d5	7.09	46	53005	38.31	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.62%
24) Chloroform-d	7.65	84	178913	22.31	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.24%
26) 1,2-Dichloroethane-d4	8.31	65	102005	23.00	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.00%
29) Benzene-d6	8.27	84	351162	23.97	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.88%
33) 1,2-Dichloropropane-d6	9.28	67	104635	23.67	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.68%
37) Toluene-d8	10.33	98	332509	22.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.80%
38) trans-1,3-Dichloropropene-	10.58	79	40298	18.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.04%
39) 2-Hexanone-d5	10.93	63	42355	39.65	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.30%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	102505	23.84	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.36%
61) 1,2-Dichlorobenzene-d4	13.86	152	118287	23.47	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.88%

Target Compounds					Ovalue
13) Acetone	4.14	43	13076	9.355 ug/L	99
64) 1,4-Dichlorobenzene	13.58	146	11278	1.240 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102618\
Data File : VW006363.D
Acq On : 25 Oct 2018 14:39
Operator : SY/AP
Sample : J5657-03
Misc : 5.96G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A40K1

Quant Time: Oct 26 04:47:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100918S.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 26 04:05:39 2018
Response via : Initial Calibration

