

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102623\
 Data File : VW027443.D
 Acq On : 26 Oct 2023 11:37
 Operator : SY/MD
 Sample : 05034-07
 Misc : 5.08g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH5L5

Quant Time: Oct 27 01:28:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 27 01:27:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	273619	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	262058	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	137225	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	53909	12.910	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.640%
7) Chloroethane-d5	2.899	69	49053	16.360	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.440%
11) 1,1-Dichloroethene-d2	4.021	65	30102	14.386	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.560%
21) 2-Butanone-d5	7.075	46	55285	39.832	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.660%
24) Chloroform-d	7.648	84	146872	18.010	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.040%
26) 1,2-Dichloroethane-d4	8.301	65	83639	19.092	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.360%
32) Benzene-d6	8.276	84	284498	16.046	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.200%
36) 1,2-Dichloropropane-d6	9.270	67	97882	18.075	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	72.280%
41) Toluene-d8	10.319	98	245275	15.255	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	61.040%
43) trans-1,3-Dichloroprop...	10.575	79	36448	14.761	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	59.040%
47) 2-Hexanone-d5	10.922	63	32437	28.950	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.900%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	80977	20.042	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.160%
66) 1,2-Dichlorobenzene-d4	13.849	152	88505	17.798	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	71.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102623\
Data File : VW027443.D
Acq On : 26 Oct 2023 11:37
Operator : SY/MD
Sample : 05034-07
Misc : 5.08g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH5L5

Quant Time: Oct 27 01:28:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 27 01:27:32 2023
Response via : Initial Calibration

