

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102824\
 Data File : VW030782.D
 Acq On : 28 Oct 2024 13:15
 Operator : SY/MD
 Sample : VW1028SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK583

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 29 01:32:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 29 01:29:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	223749	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	209804	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	92894	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	93437	26.567	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 106.280%			
7) Chloroethane-d5	2.893	69	80220	30.133	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 120.520%			
11) 1,1-Dichloroethene-d2	4.015	65	36956	25.264	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 101.040%			
21) 2-Butanone-d5	7.075	46	27102	42.820	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 85.640%			
24) Chloroform-d	7.648	84	169119	26.430	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 105.720%			
26) 1,2-Dichloroethane-d4	8.307	65	89598m	26.928	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 107.720%			
32) Benzene-d6	8.276	84	326078	26.778	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 107.120%			
36) 1,2-Dichloropropane-d6	9.270	67	95025	26.587	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 106.360%			
41) Toluene-d8	10.325	98	289762	26.067	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 104.280%			
43) trans-1,3-Dichloroprop...	10.575	79	35580	23.774	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 95.080%			
47) 2-Hexanone-d5	10.922	63	19129	41.554	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 83.100%			
56) 1,1,2,2-Tetrachloroeth...	12.690	84	61963	23.607	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 94.440%			
66) 1,2-Dichlorobenzene-d4	13.848	152	86932	27.658	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 110.640%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102824\
Data File : VW030782.D
Acq On : 28 Oct 2024 13:15
Operator : SY/MD
Sample : VW1028SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK583

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 29 01:32:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 29 01:29:12 2024
Response via : Initial Calibration

