

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
 Data File : VW030833.D
 Acq On : 30 Oct 2024 13:52
 Operator : SY/MD
 Sample : P4494-04
 Misc : 5.02g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Nov 01 01:05:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 01:03:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	312055	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	283405	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	133541	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	80769	16.467	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.880%
7) Chloroethane-d5	2.899	69	74194	19.983	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.920%
11) 1,1-Dichloroethene-d2	4.021	65	32662	16.010	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.040%
21) 2-Butanone-d5	7.075	46	38349	43.444	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.880%
24) Chloroform-d	7.654	84	190505	21.347	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.400%
26) 1,2-Dichloroethane-d4	8.307	65	98918	21.316	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.280%
32) Benzene-d6	8.276	84	325426	19.784	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.120%
36) 1,2-Dichloropropane-d6	9.270	67	107375	22.241	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.960%
41) Toluene-d8	10.325	98	300749	20.029	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.120%
43) trans-1,3-Dichloroprop...	10.575	79	40499	20.033	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.120%
47) 2-Hexanone-d5	10.922	63	28914	46.498	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	88530	24.969	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.880%
66) 1,2-Dichlorobenzene-d4	13.849	152	103115	22.821	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
Data File : VW030833.D
Acq On : 30 Oct 2024 13:52
Operator : SY/MD
Sample : P4494-04
Misc : 5.02g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Nov 01 01:05:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 01:03:00 2024
Response via : Initial Calibration

