

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
 Data File : VW030837.D
 Acq On : 30 Oct 2024 15:38
 Operator : SY/MD
 Sample : P4568-21
 Misc : 5.14g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EF7

Quant Time: Nov 01 01:06:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 01:03:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	292650	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	271424	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	135066	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	78033	16.964	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	67.840%	
7) Chloroethane-d5	2.899	69	71810	20.623	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	82.480%	
11) 1,1-Dichloroethene-d2	4.021	65	30776	16.086	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	64.360%	
21) 2-Butanone-d5	7.075	46	35733	43.165	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	86.320%	
24) Chloroform-d	7.648	84	165921	19.825	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	79.320%	
26) 1,2-Dichloroethane-d4	8.306	65	95398	21.921	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	87.680%	
32) Benzene-d6	8.276	84	310576	19.715	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	78.840%	
36) 1,2-Dichloropropane-d6	9.276	67	93706	20.266	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	81.080%	
41) Toluene-d8	10.324	98	257649	17.916	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	71.680%	
43) trans-1,3-Dichloroprop...	10.580	79	36303	18.750	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	75.000%	
47) 2-Hexanone-d5	10.922	63	27886	46.825	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	93.640%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	84596	24.913	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	99.640%	
66) 1,2-Dichlorobenzene-d4	13.854	152	94224	20.618	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	82.480%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
Data File : VW030837.D
Acq On : 30 Oct 2024 15:38
Operator : SY/MD
Sample : P4568-21
Misc : 5.14g/10mL/MSVOA_W/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EF7

Quant Time: Nov 01 01:06:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 01:03:00 2024
Response via : Initial Calibration

