

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
 Data File : VW030831.D
 Acq On : 30 Oct 2024 13:04
 Operator : SY/MD
 Sample : P4493-06
 Misc : 5.05g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Nov 01 01:04:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 01:03:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	324483	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	321030	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	124914	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	83668	16.404	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.600%
7) Chloroethane-d5	2.899	69	78597	20.358	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.440%
11) 1,1-Dichloroethene-d2	4.021	65	35909	16.927	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.720%
21) 2-Butanone-d5	7.075	46	39748	43.304	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.600%
24) Chloroform-d	7.648	84	198547	21.396	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.600%
26) 1,2-Dichloroethane-d4	8.307	65	103286	21.405	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.600%
32) Benzene-d6	8.276	84	341835	18.346	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.400%
36) 1,2-Dichloropropane-d6	9.270	67	108569	19.852	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.400%
41) Toluene-d8	10.319	98	307831	18.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	10.575	79	42252	18.451	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.800%
47) 2-Hexanone-d5	10.922	63	30172	42.835	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.660%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	85134	21.197	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.800%
66) 1,2-Dichlorobenzene-d4	13.849	152	103943	24.593	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103024\
Data File : VW030831.D
Acq On : 30 Oct 2024 13:04
Operator : SY/MD
Sample : P4493-06
Misc : 5.05g/10mL/MSVOA_W/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Nov 01 01:04:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 01:03:00 2024
Response via : Initial Calibration

