

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103123\
 Data File : VW027488.D
 Acq On : 31 Oct 2023 16:26
 Operator : SY/MD
 Sample : VIBLK562
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK562

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/01/2023
 Supervised By :Mahesh Dadoda 11/01/2023

Quant Time: Nov 01 01:50:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 27 01:27:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	443292	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	403399	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	212086	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	48767	7.208	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	28.840%#		
7) Chloroethane-d5	2.893	69	48948	10.076	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	40.320%		
11) 1,1-Dichloroethene-d2	4.021	65	29107	8.586	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	34.360%#		
21) 2-Butanone-d5	7.075	46	66018	29.359	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	58.720%		
24) Chloroform-d	7.648	84	161456m	12.220	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	48.880%		
26) 1,2-Dichloroethane-d4	8.307	65	90362	12.732	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	50.920%#		
32) Benzene-d6	8.276	84	305218	11.183	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	44.720%		
36) 1,2-Dichloropropane-d6	9.270	67	105920	12.706	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	50.840%#		
41) Toluene-d8	10.325	98	248748	10.051	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	40.200%		
43) trans-1,3-Dichloroprop...	10.581	79	33358	8.776	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	35.120%		
47) 2-Hexanone-d5	10.922	63	31469	18.246	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	36.500%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	100878	16.219	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	64.880%		
66) 1,2-Dichlorobenzene-d4	13.848	152	94328	12.273	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	49.080%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103123\
Data File : VW027488.D
Acq On : 31 Oct 2023 16:26
Operator : SY/MD
Sample : VIBLK562
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK562

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/01/2023
Supervised By : Mahesh Dadoda 11/01/2023

Quant Time: Nov 01 01:50:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 27 01:27:32 2023
Response via : Initial Calibration

