

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103124\
 Data File : VW030859.D
 Acq On : 31 Oct 2024 14:25
 Operator : SY/MD
 Sample : P4581-19
 Misc : 5.71g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 01 03:13:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 03:09:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	417576	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	406391	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	157656	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	159258	24.264	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.040%
7) Chloroethane-d5	2.899	69	130539	26.274	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.080%
11) 1,1-Dichloroethene-d2	4.021	65	63587	23.292	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	93.160%
21) 2-Butanone-d5	7.081	46	42031	35.583	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.160%
24) Chloroform-d	7.648	84	307224	25.727	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.920%
26) 1,2-Dichloroethane-d4	8.441	65	147	0.024	ug/L	0.14
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.080%#
32) Benzene-d6	8.276	84	576289	24.433	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.720%
36) 1,2-Dichloropropane-d6	9.276	67	154588	22.330	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.320%
41) Toluene-d8	10.324	98	477118	22.158	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.640%
43) trans-1,3-Dichloroprop...	10.574	79	56908	19.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.520%
47) 2-Hexanone-d5	10.922	63	33433	37.495	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.980%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	101544	19.972	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.880%
66) 1,2-Dichlorobenzene-d4	13.854	152	138089	25.886	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103124\
Data File : VW030859.D
Acq On : 31 Oct 2024 14:25
Operator : SY/MD
Sample : P4581-19
Misc : 5.71g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 01 03:13:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 03:09:59 2024
Response via : Initial Calibration

