

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103124\
 Data File : VW030863.D
 Acq On : 31 Oct 2024 16:01
 Operator : SY/MD
 Sample : P4581-23
 Misc : 4.27g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EB5

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 11/01/2024
 Supervised By :Semsettin Yesilyurt 11/04/2024

Quant Time: Nov 01 03:14:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 03:09:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	462532	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	417464	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	180779	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	146108	20.097	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	80.400%		
7) Chloroethane-d5	2.899	69	128764	23.398	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	93.600%		
11) 1,1-Dichloroethene-d2	4.021	65	56406	18.653	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	74.600%		
21) 2-Butanone-d5	7.081	46	45332	34.647	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	69.300%		
24) Chloroform-d	7.648	84	277191	20.956	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	83.840%		
26) 1,2-Dichloroethane-d4	8.301	65	147589m	21.457	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	85.840%		
32) Benzene-d6	8.276	84	528808	21.825	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	87.280%		
36) 1,2-Dichloropropane-d6	9.276	67	160239	22.532	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	90.120%		
41) Toluene-d8	10.325	98	477535	21.590	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	86.360%		
43) trans-1,3-Dichloroprop...	10.575	79	59287	19.909	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	79.640%		
47) 2-Hexanone-d5	10.922	63	33886	36.995	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	73.980%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	106839	20.456	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	81.840%		
66) 1,2-Dichlorobenzene-d4	13.848	152	145980	23.865	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	95.480%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103124\
Data File : VW030863.D
Acq On : 31 Oct 2024 16:01
Operator : SY/MD
Sample : P4581-23
Misc : 4.27g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EB5

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 11/01/2024
Supervised By :Semsettin Yesilyurt 11/04/2024

Quant Time: Nov 01 03:14:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 03:09:59 2024
Response via : Initial Calibration

