



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW103124\
Data File : VW030848.D
Acq On : 31 Oct 2024 08:56
Operator : SY/MD
Sample : BFB588
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB588

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Title : SFAM01.0
Last Update : Fri Nov 01 03:09:59 2024



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	24275	PASS
75	95	30	80	58.4	62277	PASS
95	95	100	100	100.0	106549	PASS
96	95	5	9	6.7	7113	PASS
173	174	0.00	2	0.6	479	PASS
174	95	50	120	74.5	79360	PASS
175	174	5	9	8.4	6645	PASS
176	174	95	101	99.8	79179	PASS
177	176	5	9	7.1	5611	PASS

m/z	Abundance
36.00	68.0
40.00	232.0
42.10	64.0
43.20	65.0
43.90	174.0
44.90	61.0
50.90	93.0
59.00	51.0
68.80	62.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

m/z	Abundance
36.00	1340.0
37.10	7101.0
38.10	6322.0
39.00	2479.0
40.00	326.0
41.30	121.0
41.80	76.0
42.30	76.0
42.90	233.0
44.10	1187.0
45.10	1031.0
46.10	206.0
46.90	1320.0
48.00	995.0
49.10	6181.0
50.10	24592.0
51.10	7583.0
52.00	443.0
53.00	381.0
53.90	95.0
55.00	555.0
56.00	2247.0
57.10	4127.0
58.10	200.0
59.00	155.0
60.00	1379.0
61.10	7008.0
62.10	7148.0
63.10	5798.0
64.10	869.0
65.00	1002.0
65.80	177.0
66.30	127.0
67.10	513.0
68.10	12682.0
69.10	13325.0
70.00	1342.0
71.20	95.0
72.00	626.0
73.10	6401.0
74.10	22120.0
75.10	60816.0
76.10	5042.0
77.10	615.0
78.00	458.0
79.00	5131.0
80.10	1298.0
81.00	5530.0
81.90	1180.0
82.90	146.0
86.00	222.0
87.00	4446.0
88.00	4127.0
90.00	51.0
90.90	554.0
92.00	3890.0
93.10	5244.0
94.10	14340.0
95.10	103688.0
96.00	7024.0
97.10	234.0
103.00	173.0
104.10	714.0
104.80	180.0
106.00	689.0
106.90	166.0
107.90	54.0
108.20	53.0
110.00	106.0
110.80	187.0
112.00	70.0
113.00	170.0
114.80	162.0
115.90	340.0
117.00	1042.0
117.90	511.0
119.00	882.0
119.90	62.0
123.90	97.0
125.10	194.0
126.00	152.0
127.90	335.0
128.90	144.0
129.90	349.0
130.80	191.0
132.00	53.0
133.00	633.0
133.80	127.0
134.50	97.0
135.10	305.0
136.90	169.0
140.00	97.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

140.90	862.0
141.90	156.0
142.90	1554.0
144.00	61.0
145.00	575.0
146.00	308.0
147.00	147.0
148.00	315.0
148.80	86.0
150.10	113.0
153.00	78.0
155.00	270.0
156.90	240.0
158.80	120.0
161.10	162.0
162.90	102.0
165.20	109.0
170.90	338.0
171.90	1391.0
174.00	67736.0
175.00	5785.0
176.00	64888.0
177.00	5166.0
177.90	146.0
179.00	290.0
190.00	79.0
191.00	529.0
192.00	139.0
192.90	816.0
195.10	86.0
203.20	112.0
205.00	103.0
205.90	77.0
206.30	77.0
207.20	492.0
208.20	142.0
209.00	86.0
234.90	111.0
249.00	473.0
249.90	131.0
250.90	288.0
251.80	63.0
252.90	58.0
263.70	53.0
265.10	733.0
266.10	121.0
267.20	183.0
278.30	53.0
278.80	135.0
279.30	129.0
281.10	7952.0
282.10	1623.0
283.10	1172.0
283.80	268.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

m/z	Abundance
35.20	85.0
36.00	1654.0
37.10	7639.0
38.10	6422.0
39.10	2930.0
40.20	366.0
40.80	108.0
41.80	57.0
43.20	223.0
44.00	1252.0
45.00	1449.0
46.20	228.0
47.00	1748.0
47.90	842.0
49.10	6446.0
50.10	27608.0
51.10	8470.0
52.00	376.0
53.10	206.0
54.00	148.0
55.00	529.0
56.10	2375.0
57.10	4259.0
58.10	245.0
59.20	106.0
60.10	1376.0
61.10	7366.0
62.10	7502.0
63.10	6262.0
64.10	1005.0
65.00	1020.0
65.90	127.0
67.10	453.0
68.10	14816.0
69.10	15243.0
70.10	1093.0
72.10	820.0
73.10	7428.0
74.10	24144.0
75.10	70136.0
76.10	5998.0
77.00	822.0
78.00	726.0
79.00	5957.0
80.00	1678.0
81.00	5688.0
82.00	1331.0
83.20	186.0
85.90	188.0
87.00	5372.0
88.00	5092.0
89.80	75.0
90.90	702.0
92.10	3984.0
93.10	5712.0
94.10	16504.0
95.10	121792.0
96.10	8104.0
97.10	336.0
103.00	169.0
103.90	583.0
105.00	237.0
106.00	777.0
107.00	208.0
108.20	54.0
109.80	116.0
110.20	114.0
110.90	203.0
111.90	68.0
113.00	107.0
115.00	246.0
116.00	592.0
117.00	964.0
118.00	523.0
119.00	905.0
124.00	95.0
124.80	131.0
127.90	430.0
128.90	371.0
130.00	543.0
130.90	213.0
133.10	407.0
134.20	106.0
134.90	309.0
136.00	70.0
136.90	236.0
139.80	94.0
141.00	1192.0
142.10	156.0
143.00	1982.0
143.90	114.0
145.00	513.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

146.00	215.0
146.90	212.0
147.90	258.0
148.90	115.0
149.90	186.0
151.80	102.0
152.60	66.0
155.00	323.0
155.90	79.0
157.00	236.0
158.00	54.0
158.80	155.0
159.40	57.0
160.80	124.0
163.00	81.0
165.10	90.0
169.00	72.0
170.20	114.0
171.10	296.0
172.10	1344.0
174.00	94520.0
175.00	7950.0
176.00	91936.0
177.00	6341.0
177.90	287.0
179.00	182.0
191.00	303.0
192.00	78.0
193.10	643.0
194.70	52.0
203.00	82.0
204.70	68.0
205.20	54.0
207.00	231.0
208.00	70.0
234.80	63.0
249.00	233.0
249.80	60.0
251.10	147.0
265.10	403.0
266.00	99.0
266.90	102.0
278.70	66.0
279.20	109.0
279.80	80.0
281.10	5111.0
282.10	1059.0
283.10	743.0
284.30	101.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

m/z	Abundance
35.30	81.0
36.00	885.0
37.10	4901.0
38.10	4964.0
39.10	1999.0
40.10	327.0
40.90	75.0
41.30	96.0
42.90	112.0
44.10	884.0
45.10	889.0
46.20	85.0
47.10	1052.0
48.10	821.0
49.10	4229.0
50.10	20624.0
51.10	6262.0
52.10	371.0
52.90	83.0
54.20	79.0
55.00	298.0
56.10	1703.0
57.10	2993.0
58.20	182.0
59.00	91.0
60.00	1272.0
61.00	5626.0
62.10	5235.0
63.10	4402.0
64.10	632.0
65.10	812.0
66.20	162.0
67.00	354.0
68.00	10845.0
69.10	11167.0
70.00	950.0
70.90	63.0
72.00	668.0
73.10	5194.0
74.10	18408.0
75.10	55880.0
76.10	4499.0
76.90	581.0
77.90	518.0
79.00	4337.0
80.00	1219.0
81.00	4427.0
82.00	980.0
83.00	85.0
85.90	233.0
87.00	4337.0
88.00	4068.0
91.00	623.0
92.10	3499.0
93.10	4866.0
94.10	12210.0
95.00	94168.0
96.10	6210.0
96.90	171.0
102.90	148.0
104.00	466.0
104.80	162.0
105.90	517.0
106.90	123.0
109.90	127.0
110.90	74.0
111.90	93.0
112.80	120.0
114.90	108.0
116.00	517.0
117.00	897.0
118.00	448.0
119.00	819.0
124.00	117.0
125.00	82.0
125.70	66.0
128.00	386.0
128.90	271.0
130.00	491.0
130.90	152.0
133.10	273.0
134.10	63.0
134.90	361.0
136.90	224.0
139.90	134.0
141.00	1196.0
142.00	178.0
143.00	1429.0
145.00	347.0
145.90	165.0
146.90	121.0
148.00	163.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588

148.90	124.0
149.80	78.0
151.90	84.0
152.70	111.0
154.90	265.0
155.80	73.0
157.00	214.0
158.90	193.0
160.90	137.0
163.10	75.0
165.20	63.0
168.90	75.0
170.10	65.0
171.10	203.0
172.00	1142.0
173.10	1437.0
174.00	75824.0
175.00	6199.0
176.00	80712.0
177.00	5325.0
178.00	181.0
191.00	139.0
193.00	212.0
205.00	76.0
207.20	105.0
249.00	146.0
264.60	92.0
265.10	202.0
266.10	53.0
266.90	64.0
281.10	2423.0
282.10	629.0
283.10	401.0

Instrument :
MSVOA_W
ClientSampleId :
BFB588