

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110123\
 Data File : VW027495.D
 Acq On : 01 Nov 2023 12:15
 Operator : SY/MD
 Sample : 05112-04
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Nov 02 03:34:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 02 03:33:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	365410	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	323159	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	160944	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	84021	15.066	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.280%
7) Chloroethane-d5	2.893	69	73433	18.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.360%
11) 1,1-Dichloroethene-d2	4.021	65	45217	16.182	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.720%
21) 2-Butanone-d5	7.081	46	90367	48.753	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.500%
24) Chloroform-d	7.648	84	225708	20.724	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.880%
26) 1,2-Dichloroethane-d4	8.307	65	123556	21.119	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.480%
32) Benzene-d6	8.276	84	445876	20.393	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.560%
36) 1,2-Dichloropropane-d6	9.270	67	149923	22.450	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.800%
41) Toluene-d8	10.319	98	374646	18.896	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	10.575	79	50913	16.721	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.880%
47) 2-Hexanone-d5	10.922	63	64551	46.720	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	126362	25.361	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	120140	20.599	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.400%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	16101	2.407	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110123\
Data File : VW027495.D
Acq On : 01 Nov 2023 12:15
Operator : SY/MD
Sample : 05112-04
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Nov 02 03:34:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101123SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 02 03:33:34 2023
Response via : Initial Calibration

