

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110124\
 Data File : VW030883.D
 Acq On : 01 Nov 2024 13:38
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 04 00:54:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 03:09:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	344119	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	330678	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	148442	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	104748	19.365	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	77.480%	
7) Chloroethane-d5	2.899	69	95798	23.397	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	93.600%	
11) 1,1-Dichloroethene-d2	4.021	65	43473	19.323	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	77.280%	
21) 2-Butanone-d5	7.081	46	54414	55.900	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	111.800%	
24) Chloroform-d	7.648	84	246519	25.050	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	100.200%	
26) 1,2-Dichloroethane-d4	8.300	65	134958	26.372	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	105.480%	
32) Benzene-d6	8.276	84	430835	22.448	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	89.800%	
36) 1,2-Dichloropropane-d6	9.276	67	139897	24.834	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	99.320%	
41) Toluene-d8	10.325	98	381514	21.775	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	87.120%	
43) trans-1,3-Dichloroprop...	10.574	79	54852	23.254	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	93.000%	
47) 2-Hexanone-d5	10.922	63	42082	58.000	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	116.000%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	114226	27.611	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	110.440%	
66) 1,2-Dichlorobenzene-d4	13.848	152	138578	27.591	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	110.360%	
Target Compounds						
39) cis-1,3-Dichloropropene	10.032	75	4173	0.515	ug/L	# 63
71) Naphthalene	15.360	128	6226	0.678	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110124\
Data File : VW030883.D
Acq On : 01 Nov 2024 13:38
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 04 00:54:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 03:09:59 2024
Response via : Initial Calibration

