

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110124\
 Data File : VW030885.D
 Acq On : 01 Nov 2024 14:59
 Operator : SY/MD
 Sample : P4585-06
 Misc : 5.27g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 04 00:54:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 01 03:09:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	388715	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	364057	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	158739	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	108535	17.763	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.040%
7) Chloroethane-d5	2.899	69	93287	20.170	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.680%
11) 1,1-Dichloroethene-d2	4.021	65	46869	18.443	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.760%
21) 2-Butanone-d5	7.075	46	52177	47.452	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.900%
24) Chloroform-d	7.648	84	233550	21.009	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.040%
26) 1,2-Dichloroethane-d4	8.301	65	128688	22.262	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.040%
32) Benzene-d6	8.270	84	420198	19.886	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.560%
36) 1,2-Dichloropropane-d6	9.270	67	133363	21.504	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	86.000%
41) Toluene-d8	10.325	98	369393	19.150	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.600%
43) trans-1,3-Dichloroprop...	10.581	79	51685	19.902	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.600%
47) 2-Hexanone-d5	10.922	63	39922	49.978	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.960%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	108667	23.859	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.440%
66) 1,2-Dichlorobenzene-d4	13.849	152	131527	24.488	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110124\
Data File : VW030885.D
Acq On : 01 Nov 2024 14:59
Operator : SY/MD
Sample : P4585-06
Misc : 5.27g/10mL/MSVOA_W/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 04 00:54:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 01 03:09:59 2024
Response via : Initial Calibration

