

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110222\
 Data File : VW025012.D
 Acq On : 02 Nov 2022 17:06
 Operator : SY/VA
 Sample : VSTDCCC025EC
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTD025484

Quant Time: Nov 03 06:31:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 27 23:51:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	279204	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	255655	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	135019	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	60359	19.260	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.040%
7) Chloroethane-d5	2.885	69	38265	20.844	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.360%
11) 1,1-Dichloroethene-d2	4.019	63	166371	21.312	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.240%
21) 2-Butanone-d5	7.079	46	46951	44.390	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.780%
24) Chloroform-d	7.646	84	212092	25.030	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.120%
26) 1,2-Dichloroethane-d4	8.305	65	120784	25.465	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.880%
32) Benzene-d6	8.274	84	399475	24.378	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.520%
36) 1,2-Dichloropropane-d6	9.274	67	114078	23.759	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.040%
41) Toluene-d8	10.323	98	368398	24.148	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	10.573	79	51624	24.248	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.000%
47) 2-Hexanone-d5	10.920	63	38271	46.143	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.280%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	91096	23.762	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	125268	24.142	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.560%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.007	85	23411	25.139	ug/L	89
3) Chloromethane	2.214	50	42121	18.805	ug/L	100
5) Vinyl chloride	2.361	62	65311	21.103	ug/L	97
6) Bromomethane	2.775	94	33107	21.436	ug/L	98
8) Chloroethane	2.922	64	26485	20.515	ug/L	92
9) Trichlorofluoromethane	3.257	101	56225	21.656	ug/L	99
10) 1,1,2-Trichloro-1,2,2-...	4.062	101	87849	23.895	ug/L	99
12) 1,1-Dichloroethene	4.037	96	81919	23.653	ug/L	95
13) Acetone	4.129	43	27704	36.871	ug/L	97
14) Carbon disulfide	4.385	76	228173	22.553	ug/L	97
15) Methyl Acetate	4.677	43	33641	22.488	ug/L	99
16) Methylene chloride	4.921	84	96016	19.368	ug/L	99
17) trans-1,2-Dichloroethene	5.427	96	89500	24.133	ug/L	96
18) Methyl tert-butyl Ether	5.427	73	155220	24.951	ug/L	99
19) 1,1-Dichloroethane	6.220	63	167866	24.449	ug/L	99
20) cis-1,2-Dichloroethene	7.171	96	101044	24.998	ug/L	99
22) 2-Butanone	7.171	43	43703	38.821	ug/L	97
23) Bromochloromethane	7.518	128	43917	25.062	ug/L	96
25) Chloroform	7.677	83	188164	25.967	ug/L	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110222\
 Data File : VW025012.D
 Acq On : 02 Nov 2022 17:06
 Operator : SY/VA
 Sample : VSTDCCC025EC
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTD025484

Quant Time: Nov 03 06:31:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 27 23:51:01 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
27) 1,2-Dichloroethane	8.402	62	128327	26.108	ug/L	98
29) Cyclohexane	7.957	56	136528	22.617	ug/L	97
30) 1,1,1-Trichloroethane	7.872	97	160822	25.706	ug/L	98
31) Carbon tetrachloride	8.067	117	143174	25.610	ug/L	99
33) Benzene	8.323	78	380454	24.712	ug/L	100
34) Trichloroethene	9.091	95	101049	24.794	ug/L	96
35) Methylcyclohexane	9.335	83	162155	23.416	ug/L	100
37) 1,2-Dichloropropane	9.372	63	93054	24.903	ug/L	99
38) Bromodichloromethane	9.646	83	139510	26.751	ug/L	97
39) cis-1,3-Dichloropropene	10.073	75	151255	25.605	ug/L	98
40) 4-Methyl-2-pentanone	10.207	43	101488	46.284	ug/L	98
42) Toluene	10.390	91	410750	25.131	ug/L	96
44) trans-1,3-Dichloropropene	10.603	75	132888	25.304	ug/L	98
45) 1,1,2-Trichloroethane	10.786	97	70453	24.972	ug/L	99
46) Tetrachloroethene	10.859	164	70803	24.292	ug/L	92
48) 2-Hexanone	10.969	43	69046	45.531	ug/L	98
49) Dibromochloromethane	11.128	129	90786	26.399	ug/L	99
50) 1,2-Dibromoethane	11.231	107	66583	24.954	ug/L	92
51) Chlorobenzene	11.658	112	266198	25.126	ug/L	99
52) Ethylbenzene	11.725	91	470251	24.925	ug/L	97
53) m,p-Xylene	11.835	106	182136	24.940	ug/L	100
54) o-Xylene	12.164	106	178563	25.917	ug/L	98
55) Styrene	12.176	104	310002	25.906	ug/L	93
57) 1,1,2,2-Tetrachloroethane	12.713	83	83094	24.551	ug/L	99
59) Bromoform	12.347	173	46399	24.297	ug/L	94
60) Isopropylbenzene	12.463	105	487736	24.027	ug/L	100
61) 1,2,3-Trichloropropane	12.767	75	62095	22.891	ug/L	97
62) 1,3,5-Trimethylbenzene	12.938	105	436796	24.788	ug/L	100
63) 1,2,4-Trimethylbenzene	13.249	105	427050	24.714	ug/L	99
64) 1,3-Dichlorobenzene	13.493	146	206489	24.389	ug/L	93
65) 1,4-Dichlorobenzene	13.578	146	207446	24.556	ug/L	97
67) 1,2-Dichlorobenzene	13.865	146	186496	24.499	ug/L #	87
68) 1,2-Dibromo-3-chloropr...	14.481	75	14702	23.453	ug/L	96
69) 1,3,5-Trichlorobenzene	14.627	180	139918	24.951	ug/L	99
70) 1,2,4-trichlorobenzene	15.127	180	105705	24.253	ug/L	98
71) Naphthalene	15.358	128	227069	24.696	ug/L	99
72) 1,2,3-Trichlorobenzene	15.547	180	92171	24.271	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110222\
Data File : VW025012.D
Acq On : 02 Nov 2022 17:06
Operator : SY/VA
Sample : VSTDCCC025EC
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTD025484

Quant Time: Nov 03 06:31:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 27 23:51:01 2022
Response via : Initial Calibration









#2
 Dichlorodifluoromethane
 Concen: 25.139 ug/L
 RT: 2.007 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VW025012.D
 Acq: 02 Nov 2022 17:06




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTD025484


Tgt Ion: 85 Resp: 23411
 Ion Ratio Lower Upper
 85 100
 87 26.3 25.9 38.9



#3
 Chloromethane
 Concen: 18.805 ug/L
 RT: 2.214 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VW025012.D
 Acq: 02 Nov 2022 17:06

Tgt Ion: 50 Resp: 42121
 Ion Ratio Lower Upper
 50 100
 52 34.6 24.1 44.8









#5
 Vinyl chloride
 Concen: 21.103 ug/L
 RT: 2.361 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: VW025012.D
 Acq: 02 Nov 2022 17:06




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTD025484


Tgt Ion: 62 Resp: 65311
 Ion Ratio Lower Upper
 62 100
 64 37.4 27.2 50.6



#6
 Bromomethane
 Concen: 21.436 ug/L
 RT: 2.775 min Scan# 143
 Delta R.T. -0.006 min
 Lab File: VW025012.D
 Acq: 02 Nov 2022 17:06

Tgt Ion: 94 Resp: 33107
 Ion Ratio Lower Upper
 94 100
 96 93.1 63.8 118.6









#8
Chloroethane
Concen: 20.515 ug/L
RT: 2.922 min Scan# 1
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 64 Resp: 26485
Ion Ratio Lower Upper
64 100
66 35.1 21.6 40.2



#9
Trichlorofluoromethane
Concen: 21.656 ug/L
RT: 3.257 min Scan# 222
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 101 Resp: 56225
Ion Ratio Lower Upper
101 100
103 63.2 51.4 77.0









#10
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 23.895 ug/L
RT: 4.062 min Scan# 310
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:101 Resp: 87849
Ion Ratio Lower Upper
101 100
85 44.7 36.5 54.7
151 70.5 57.3 85.9



#12
1,1-Dichloroethene
Concen: 23.653 ug/L
RT: 4.037 min Scan# 350
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 96 Resp: 81919
Ion Ratio Lower Upper
96 100
61 176.2 126.2 234.4
63 150.4 98.3 182.7





#13

Acetone

Concen: 36.871 ug/L

RT: 4.129 min Scan# 365

Delta R.T. -0.006 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Instrument :

MSVOA_W

ClientSampleId :





VSTD025484

Tgt Ion: 43 Resp: 27704

Ion Ratio Lower Upper

43 100

58 31.0 0.0 65.4



#14

Carbon disulfide

Concen: 22.553 ug/L

RT: 4.385 min Scan# 407

Delta R.T. -0.006 min

Lab File: VW025012.D




Acq: 02 Nov 2022 17:06


Tgt Ion: 76 Resp: 228173

Ion Ratio Lower Upper

76 100

78 9.6 6.8 10.2









#15
Methyl Acetate
Concen: 22.488 ug/L
RT: 4.677 min Scan# 41
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 43 Resp: 33641
Ion Ratio Lower Upper
43 100
74 26.2 20.6 30.8



#16
Methylene chloride
Concen: 19.368 ug/L
RT: 4.921 min Scan# 495
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 84 Resp: 96016
Ion Ratio Lower Upper
84 100
86 61.9 42.6 79.0
49 116.3 81.9 152.1









#17
trans-1,2-Dichloroethene
Concen: 24.133 ug/L
RT: 5.427 min Scan# 577
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 96 Resp: 89500
Ion Ratio Lower Upper
96 100
61 144.6 97.6 181.2
98 63.6 46.8 86.8



#18
Methyl tert-butyl Ether
Concen: 24.951 ug/L
RT: 5.427 min Scan# 578
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 73 Resp: 155220
Ion Ratio Lower Upper
73 100
43 21.0 16.6 24.8
57 20.5 17.1 25.7





#19

1,1-Dichloroethane

Concen: 24.449 ug/L

RT: 6.220 min Scan# 707

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Instrument :

MSVOA_W

ClientSampleId :

VSTD025484





Tgt Ion: 63 Resp: 167866

Ion Ratio Lower Upper

63 100

65 32.2 22.2 41.2

83 13.8 9.7 17.9



#20

cis-1,2-Dichloroethene

Concen: 24.998 ug/L

RT: 7.171 min Scan# 864

Delta R.T. 0.006 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06




Tgt Ion: 96 Resp: 101044


Ion Ratio Lower Upper

96 100

61 130.2 91.8 170.6

68 0.0 0.0 0.0









#22
2-Butanone
Concen: 38.821 ug/L
RT: 7.171 min Scan# 864
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 43 Resp: 43703
Ion Ratio Lower Upper
43 100
72 31.0 14.6 44.0



#23
Bromochloromethane
Concen: 25.062 ug/L
RT: 7.518 min Scan# 921
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 128 Resp: 43917
Ion Ratio Lower Upper
128 100
49 154.9 112.8 209.4
130 126.3 86.2 160.2
51 49.9 41.6 62.4









#25
Chloroform
Concen: 25.967 ug/L
RT: 7.677 min Scan# 94
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 83 Resp: 188164
Ion Ratio Lower Upper
83 100
85 63.5 47.5 88.3



#27
1,2-Dichloroethane
Concen: 26.108 ug/L
RT: 8.402 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 62 Resp: 128327
Ion Ratio Lower Upper
62 100
98 8.0 7.0 10.4









#29
Cyclohexane
Concen: 22.617 ug/L
RT: 7.957 min Scan# 993
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 56 Resp: 136528
Ion Ratio Lower Upper
56 100
69 32.1 25.0 37.4
84 94.3 72.6 108.8



#30
1,1,1-Trichloroethane
Concen: 25.706 ug/L
RT: 7.872 min Scan# 979
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 97 Resp: 160822
Ion Ratio Lower Upper
97 100
99 63.6 51.8 77.8
61 44.4 36.6 54.8









#31
Carbon tetrachloride
Concen: 25.610 ug/L
RT: 8.067 min Scan# 1011
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:117 Resp: 143174
Ion Ratio Lower Upper
117 100
119 98.1 77.7 116.5



#33
Benzene
Concen: 24.712 ug/L
RT: 8.323 min Scan# 1053
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 78 Resp: 380454









#34
Trichloroethene
Concen: 24.794 ug/L
RT: 9.091 min Scan# 1179
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 95 Resp: 101049
Ion Ratio Lower Upper
95 100
97 65.2 44.7 83.1
132 97.0 62.6 116.2
130 100.1 69.1 128.3



#35
Methylcyclohexane
Concen: 23.416 ug/L
RT: 9.335 min Scan# 1219
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 83 Resp: 162155
Ion Ratio Lower Upper
83 100
55 75.1 60.4 90.6
98 49.3 39.5 59.3





#37

1,2-Dichloropropane

Concen: 24.903 ug/L

RT: 9.372 min Scan# 11

Delta R.T. 0.006 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Instrument :

MSVOA_W

ClientSampleId :





VSTD025484

Tgt Ion: 63 Resp: 93054

Ion Ratio Lower Upper

63 100

112 4.8 3.6 5.4



#38

Bromodichloromethane

Concen: 26.751 ug/L

RT: 9.646 min Scan# 1270

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06




Tgt Ion: 83 Resp: 139510


Ion Ratio Lower Upper

83 100

85 62.9 45.8 85.0

127 8.8 6.8 10.2





#39

cis-1,3-Dichloropropene

Concen: 25.605 ug/L

RT: 10.073 min Scan# 11

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Instrument :

MSVOA_W

ClientSampleId :





VSTD025484

Tgt Ion: 75 Resp: 151255

Ion Ratio Lower Upper

75 100

77 31.0 22.6 42.0



#40

4-Methyl-2-pentanone

Concen: 46.284 ug/L

RT: 10.207 min Scan# 1362

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06




Tgt Ion: 43 Resp: 101488


Ion Ratio Lower Upper

43 100

58 41.6 32.6 48.8

100 18.6 14.0 21.0









#42
Toluene
Concen: 25.131 ug/L
RT: 10.390 min Scan# 1392
Delta R.T. 0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 91 Resp: 410750
Ion Ratio Lower Upper
91 100
92 58.3 38.6 71.8



#44
trans-1,3-Dichloropropene
Concen: 25.304 ug/L
RT: 10.603 min Scan# 1427
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 75 Resp: 132888
Ion Ratio Lower Upper
75 100
77 33.1 22.3 41.5





#45

1,1,2-Trichloroethane

Concen: 24.972 ug/L

RT: 10.786 min Scan# 1457

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Instrument :

MSVOA_W

ClientSampleId :

VSTD025484

Tgt Ion: 97 Resp: 70453





Ion Ratio Lower Upper

97 100

99 64.4 44.0 81.6

83 87.1 59.9 111.3

85 54.2 37.7 69.9



#46

Tetrachloroethene

Concen: 24.292 ug/L

RT: 10.859 min Scan# 1469

Delta R.T. -0.000 min

Lab File: VW025012.D

Acq: 02 Nov 2022 17:06

Tgt Ion:164 Resp: 70803




Ion Ratio Lower Upper


164 100

129 97.7 72.9 135.5

131 91.3 70.1 130.3

166 125.1 80.2 148.9









#48
2-Hexanone
Concen: 45.531 ug/L
RT: 10.969 min Scan# 1486
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:	43	Resp:	69046
Ion	Ratio	Lower	Upper
43	100		
58	56.3	46.1	69.1
57	21.3	16.2	24.2
100	14.8	11.0	16.6



#49
Dibromochloromethane
Concen: 26.399 ug/L
RT: 11.128 min Scan# 1513
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:	129	Resp:	90786
Ion	Ratio	Lower	Upper
129	100		
127	76.7	53.2	98.8









#50
1,2-Dibromoethane
Concen: 24.954 ug/L
RT: 11.231 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Instrument :
MSVOA_W
ClientSampleId :
VSTD025484

Tgt Ion:107 Resp: 66583




Ion	Ratio	Lower	Upper
107	100		
109	99.6	64.5	119.7
188	3.3	2.9	4.3




#51
Chlorobenzene
Concen: 25.126 ug/L
RT: 11.658 min Scan# 1600
Delta R.T. 0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:112 Resp: 266198

Ion	Ratio	Lower	Upper
112	100		
114	30.7	20.8	38.6
77	60.5	48.2	72.2









#52
Ethylbenzene
Concen: 24.925 ug/L
RT: 11.725 min Scan# 1611
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion: 91 Resp: 470251
Ion Ratio Lower Upper
91 100
106 30.2 22.3 41.5



#53
m,p-Xylene
Concen: 24.940 ug/L
RT: 11.835 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:106 Resp: 182136
Ion Ratio Lower Upper
106 100
91 205.4 143.9 267.2









#54
o-Xylene
Concen: 25.917 ug/L
RT: 12.164 min Scan# 1
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:106 Resp: 178563
Ion Ratio Lower Upper
106 100
91 222.6 153.5 285.1



#55
Styrene
Concen: 25.906 ug/L
RT: 12.176 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:104 Resp: 310002
Ion Ratio Lower Upper
104 100
78 49.0 31.3 58.1









#57
1,1,2,2-Tetrachloroethane
Concen: 24.551 ug/L
RT: 12.713 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:	83	Resp:	83094
Ion	Ratio	Lower	Upper
83	100		
85	67.5	46.4	86.2
131	10.5	8.2	12.4



#59
Bromoform
Concen: 24.297 ug/L
RT: 12.347 min Scan# 1713
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:	173	Resp:	46399
Ion	Ratio	Lower	Upper
173	100		
175	51.4	37.4	56.2
254	7.8	5.8	8.8









#60
Isopropylbenzene
Concen: 24.027 ug/L
RT: 12.463 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:105 Resp: 487736
Ion Ratio Lower Upper
105 100
120 25.7 20.8 31.2
77 16.1 13.0 19.4



#61
1,2,3-Trichloropropane
Concen: 22.891 ug/L
RT: 12.767 min Scan# 1782
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 75 Resp: 62095
Ion Ratio Lower Upper
75 100
77 32.6 25.7 38.5
110 42.0 31.7 47.5









#62
1,3,5-Trimethylbenzene
Concen: 24.788 ug/L
RT: 12.938 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:105 Resp: 436796
Ion Ratio Lower Upper
105 100
120 47.1 37.6 56.4



#63
1,2,4-Trimethylbenzene
Concen: 24.714 ug/L
RT: 13.249 min Scan# 1861
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:105 Resp: 427050
Ion Ratio Lower Upper
105 100
120 43.8 35.7 53.5









#64
1,3-Dichlorobenzene
Concen: 24.389 ug/L
RT: 13.493 min Scan# 1901
Delta R.T. -0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Instrument :
MSVOA_W
ClientSampleId :
VSTD025484

Tgt Ion:146 Resp: 206489




Ion	Ratio	Lower	Upper
146	100		
111	44.5	31.1	57.7
148	58.3	47.2	87.6




#65
1,4-Dichlorobenzene
Concen: 24.556 ug/L
RT: 13.578 min Scan# 1915
Delta R.T. 0.006 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:146 Resp: 207446

Ion	Ratio	Lower	Upper
146	100		
111	39.3	28.8	53.6
148	62.0	44.7	82.9









#67
1,2-Dichlorobenzene
Concen: 24.499 ug/L
RT: 13.865 min Scan# 1962
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06




Instrument :
MSVOA_W
ClientSampleId :
VSTD025484

Tgt Ion:146 Resp: 186496
Ion Ratio Lower Upper
146 100
111 48.5 34.2 51.4
148 71.0 47.1 70.7#




#68
1,2-Dibromo-3-chloropropane
Concen: 23.453 ug/L
RT: 14.481 min Scan# 2063
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion: 75 Resp: 14702
Ion Ratio Lower Upper
75 100
155 75.7 61.2 91.8
157 96.8 72.4 108.6




Abundance Scan 2086 (14.621 min): VW024999.D\data.ms (




m/z-->

Abundance Scan 2087 (14.627 min): VW025012.D\data.ms




m/z-->

Abundance Scan 2087 (14.627 min): VW025012.D\data.ms (




m/z-->

Abundance Scan 2169 (15.127 min): VW024999.D\data.ms (




m/z-->

Abundance Scan 2169 (15.127 min): VW025012.D\data.ms



m/z-->

Abundance Scan 2169 (15.127 min): VW025012.D\data.ms (




m/z-->

#69
1,3,5-Trichlorobenzene
Concen: 24.951 ug/L
RT: 14.627 min Scan# 2087
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06


Instrument :
MSVOA_W
ClientSampleId :
VSTD025484


Tgt Ion:180	Resp:	139918
Ion	Ratio	Lower Upper
180	100	
182	95.1	75.7 113.5
184	28.2	23.4 35.2
145	32.4	27.3 40.9



#70
1,2,4-trichlorobenzene
Concen: 24.253 ug/L
RT: 15.127 min Scan# 2169
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:180	Resp:	105705
Ion	Ratio	Lower Upper
180	100	
182	95.5	74.9 112.3
145	35.2	27.0 40.4









#71
Naphthalene
Concen: 24.696 ug/L
RT: 15.358 min Scan# 21
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Instrument :
MSVOA_W
ClientSampleId :
VSTD025484

Tgt Ion:	128	Resp:	227069
Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.4	15.6
129	11.1	8.6	13.0



#72
1,2,3-Trichlorobenzene
Concen: 24.271 ug/L
RT: 15.547 min Scan# 2238
Delta R.T. -0.000 min
Lab File: VW025012.D
Acq: 02 Nov 2022 17:06

Tgt Ion:	180	Resp:	92171
Ion	Ratio	Lower	Upper
180	100		
182	95.5	74.1	111.1
145	37.0	28.2	42.2

