

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110222\
 Data File : VW025009.D
 Acq On : 02 Nov 2022 15:53
 Operator : SY/VA
 Sample : N5421-01RE
 Misc : 3.33g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C1113RE

Quant Time: Nov 03 06:30:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 27 23:51:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	235308	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	135887	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	25260	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	42361	16.039	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.160%
7) Chloroethane-d5	2.885	69	29227	18.890	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.560%
11) 1,1-Dichloroethene-d2	4.013	63	96433	14.657	ug/L	-0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.640%
21) 2-Butanone-d5	7.085	46	36438	40.877	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.760%
24) Chloroform-d	7.652	84	176099	24.660	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.640%
26) 1,2-Dichloroethane-d4	8.311	65	94137	23.550	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	8.274	84	277777	31.892	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	127.560%
36) 1,2-Dichloropropane-d6	9.274	67	96093	37.653	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	150.600%#
41) Toluene-d8	10.323	98	215911	26.627	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	106.520%
43) trans-1,3-Dichloroprop...	10.579	79	29548	26.111	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.440%
47) 2-Hexanone-d5	10.926	63	27384	62.117	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	67999	33.370	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	133.480%#
66) 1,2-Dichlorobenzene-d4	13.852	152	22090	22.756	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.040%
Target Compounds						Qvalue
34) Trichloroethene	9.091	95	2854	1.318	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110222\
Data File : VW025009.D
Acq On : 02 Nov 2022 15:53
Operator : SY/VA
Sample : N5421-01RE
Misc : 3.33g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C1113RE

Quant Time: Nov 03 06:30:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 27 23:51:01 2022
Response via : Initial Calibration

