

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW110619\
 Data File : VW013918.D
 Acq On : 06 Nov 2019 11:38
 Operator : SY/VA
 Sample : K5698-03
 Misc : 7.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 H0AC4

Quant Time: Nov 07 00:56:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110119S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 07 00:53:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	934	25.00	ug/L	0.01
28) Chlorobenzene-d5	11.63	117	270	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	116	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	361	24.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.36%
7) Chloroethane-d5	2.89	69	226	19.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.84%
10) 1,1-Dichloroethene-d2	4.04	63	422	14.26	ug/L	0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.04%
20) 2-Butanone-d5	7.09	46	53	10.13	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	20.26%
24) Chloroform-d	7.65	84	1117	37.96	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	151.84%#
26) 1,2-Dichloroethane-d4	8.29	65	195	12.31	ug/L	-0.01
Spiked Amount	25.000	Range	70 - 130	Recovery	=	49.24%#
29) Benzene-d6	8.27	84	1657	91.61	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	366.44%#
33) 1,2-Dichloropropane-d6	9.27	67	501	91.50	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	366.00%#
37) Toluene-d8	10.32	98	486	28.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	113.28%
38) trans-1,3-Dichloropropene-	0.00	79	0	0.00	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.00%#
39) 2-Hexanone-d5	0.00	63	0	0.00	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.00%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	309	63.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	255.20%#
61) 1,2-Dichlorobenzene-d4	13.57	152	116	25.27	ug/L	-0.29
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.08%

Target Compounds

					Qvalue
13) Acetone	4.14	43	112	21.990 ug/L #	43
15) Methyl Acetate	4.73	43	25	2.835 ug/L #	51
22) cis-1,2-Dichloroethene	7.20	96	42	2.377 ug/L #	2
30) Cyclohexane	8.25	56	44	5.400 ug/L #	15
34) Benzene	8.43	78	131	6.683 ug/L	100
36) Methylcyclohexane	9.27	83	85	9.372 ug/L #	18
40) 1,2-Dichloropropane	9.27	63	57	11.847 ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW110619\
Data File : VW013918.D
Acq On : 06 Nov 2019 11:38
Operator : SY/VA
Sample : K5698-03
Misc : 7.01G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
H0AC4

Quant Time: Nov 07 00:56:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110119S.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 07 00:53:52 2019
Response via : Initial Calibration

