

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110621\
 Data File : VW020793.D
 Acq On : 05 Nov 2021 20:49
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK476

Quant Time: Nov 08 04:55:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 08 04:38:07 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	133687	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.634	117	120900	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.560	152	65821	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.349	65	38426	30.571	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 122.280%		
7) Chloroethane-d5		2.879	69	18968	26.637	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.560%		
11) 1,1-Dichloroethene-d2		4.013	63	47670	18.068	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 72.280%		
21) 2-Butanone-d5		7.080	46	24347	58.325	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 116.640%		
24) Chloroform-d		7.647	84	73775	24.836	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 99.360%		
26) 1,2-Dichloroethane-d4		8.305	65	42618	26.951	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 107.800%		
32) Benzene-d6		8.275	84	149661	27.399	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 109.600%		
36) 1,2-Dichloropropane-d6		9.275	67	44451	26.284	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 105.120%		
41) Toluene-d8		10.323	98	147075	27.676	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 110.720%		
43) trans-1,3-Dichloroprop...		10.579	79	22075	26.317	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 105.280%		
47) 2-Hexanone-d5		10.921	63	19994	58.430	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 116.860%		
56) 1,1,2,2-Tetrachloroeth...		12.689	84	39190	27.961	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 111.840%		
66) 1,2-Dichlorobenzene-d4		13.853	152	56225	27.181	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 108.720%		
Target Compounds							Qvalue
34) Trichloroethene		9.098	95	1140	0.591	ug/L	95
42) Toluene		10.390	91	13863	1.780	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110621\
Data File : VW020793.D
Acq On : 05 Nov 2021 20:49
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK476

Quant Time: Nov 08 04:55:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 08 04:38:07 2021
Response via : Initial Calibration

