

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110621\
 Data File : VW020786.D
 Acq On : 05 Nov 2021 18:00
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK474

Quant Time: Nov 08 04:53:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 08 04:38:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	138738	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	125565	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	70469	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	39447	30.241	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 120.960%	
7) Chloroethane-d5	2.879	69	21796	29.495	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 117.960%	
11) 1,1-Dichloroethene-d2	4.013	63	49457	18.063	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 72.240%	
21) 2-Butanone-d5	7.074	46	24674	56.956	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 113.920%	
24) Chloroform-d	7.653	84	74886	24.292	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 97.160%	
26) 1,2-Dichloroethane-d4	8.305	65	43174	26.308	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 105.240%	
32) Benzene-d6	8.275	84	151789	26.756	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 107.040%	
36) 1,2-Dichloropropane-d6	9.274	67	44770	25.489	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 101.960%	
41) Toluene-d8	10.323	98	146081	26.467	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 105.880%	
43) trans-1,3-Dichloroprop...	10.579	79	22913	26.301	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 105.200%	
47) 2-Hexanone-d5	10.921	63	20859	58.693	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 117.380%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	39578	27.189	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 108.760%	
66) 1,2-Dichlorobenzene-d4	13.853	152	57164	25.812	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 103.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110621\
Data File : VW020786.D
Acq On : 05 Nov 2021 18:00
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK474

Quant Time: Nov 08 04:53:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 08 04:38:07 2021
Response via : Initial Calibration

