

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110822\
 Data File : VW025022.D
 Acq On : 08 Nov 2022 12:50
 Operator : SY/VA
 Sample : N5519-01
 Misc : 6.14g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C1119

Quant Time: Nov 09 00:56:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Nov 09 00:55:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	290585	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	266866	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	138815	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	38078	11.675	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	46.680%
7) Chloroethane-d5	2.885	69	28107	14.711	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.840%
11) 1,1-Dichloroethene-d2	4.007	63	88230	10.860	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	43.440%#
21) 2-Butanone-d5	7.074	46	40068	36.399	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.800%
24) Chloroform-d	7.647	84	160134	18.158	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.640%
26) 1,2-Dichloroethane-d4	8.305	65	94049	19.052	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.200%
32) Benzene-d6	8.275	84	288647	16.875	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	67.480%
36) 1,2-Dichloropropane-d6	9.274	67	86477	17.254	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	69.000%#
41) Toluene-d8	10.317	98	268624	16.868	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.480%
43) trans-1,3-Dichloroprop...	10.579	79	40751	18.337	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.360%
47) 2-Hexanone-d5	10.920	63	33592	38.800	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.600%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	73647	18.403	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	73.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	98840	18.528	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.120%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110822\
Data File : VW025022.D
Acq On : 08 Nov 2022 12:50
Operator : SY/VA
Sample : N5519-01
Misc : 6.14g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C1119

Quant Time: Nov 09 00:56:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Nov 09 00:55:34 2022
Response via : Initial Calibration

