

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110824\
 Data File : VW030968.D
 Acq On : 08 Nov 2024 14:27
 Operator : SY/MD
 Sample : P4751-06
 Misc : 5.66g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9E3

Quant Time: Nov 08 22:06:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 08 22:04:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	28688	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	26576	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.562	152	9280	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	10533	22.386	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.560%
7) Chloroethane-d5	2.893	69	9745	27.005	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.000%
11) 1,1-Dichloroethene-d2	4.027	65	4616	23.776	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	95.120%
21) 2-Butanone-d5	7.093	46	4261	55.031	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.060%
24) Chloroform-d	7.648	84	21908	25.904	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	8.306	65	11391	27.089	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.360%
32) Benzene-d6	8.276	84	37825	22.910	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.640%
36) 1,2-Dichloropropane-d6	9.276	67	12665	26.022	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.080%
41) Toluene-d8	10.324	98	26232	17.713	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.840%
43) trans-1,3-Dichloroprop...	10.587	79	3790	19.211	ug/L	0.01
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.840%
47) 2-Hexanone-d5	10.928	63	2314	36.685	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.360%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	12221	36.040	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	144.160%#
66) 1,2-Dichlorobenzene-d4	13.848	152	8068	23.867	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110824\
Data File : VW030968.D
Acq On : 08 Nov 2024 14:27
Operator : SY/MD
Sample : P4751-06
Misc : 5.66g/10mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9E3

Quant Time: Nov 08 22:06:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 08 22:04:34 2024
Response via : Initial Calibration

