

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110824\
 Data File : VW030972.D
 Acq On : 08 Nov 2024 16:03
 Operator : SY/MD
 Sample : P4724-15
 Misc : 6.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9F0

Quant Time: Nov 08 22:07:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 08 22:04:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	534637	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	456893	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	165069	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	136495	15.566	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.280%
7) Chloroethane-d5	2.905	69	105947	15.754	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.000%
11) 1,1-Dichloroethene-d2	4.027	65	60075	16.604	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.400%
21) 2-Butanone-d5	7.075	46	52966	36.706	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.420%
24) Chloroform-d	7.648	84	288496	18.304	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.200%
26) 1,2-Dichloroethane-d4	8.307	65	143821	18.353	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.400%
32) Benzene-d6	8.276	84	548892	19.338	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.360%
36) 1,2-Dichloropropane-d6	9.276	67	172019	20.558	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.240%
41) Toluene-d8	10.325	98	456819	17.943	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.760%
43) trans-1,3-Dichloroprop...	10.581	79	58822	17.343	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.360%
47) 2-Hexanone-d5	10.922	63	41290	38.075	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.160%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	123575	21.197	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.800%
66) 1,2-Dichlorobenzene-d4	13.848	152	122542	20.380	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110824\
Data File : VW030972.D
Acq On : 08 Nov 2024 16:03
Operator : SY/MD
Sample : P4724-15
Misc : 6.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9F0

Quant Time: Nov 08 22:07:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 08 22:04:34 2024
Response via : Initial Calibration

