

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111122\
 Data File : VW025040.D
 Acq On : 11 Nov 2022 17:21
 Operator : SY/VA
 Sample : N5603-05
 Misc : 5.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AH6

Quant Time: Nov 12 00:49:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 12 00:48:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	23319	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	22603	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	8057	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	3371	12.879	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.520%
7) Chloroethane-d5	2.891	69	4585	29.904	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.600%
11) 1,1-Dichloroethene-d2	4.031	63	7760	11.902	ug/L	0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	47.600%
21) 2-Butanone-d5	7.092	46	3107m	35.172	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.340%
24) Chloroform-d	7.659	84	13674m	19.322	ug/L	0.01
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.280%
26) 1,2-Dichloroethane-d4	8.311	65	8288	20.922	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.680%
32) Benzene-d6	8.281	84	24316	16.784	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	67.120%
36) 1,2-Dichloropropane-d6	9.281	67	7274	17.136	ug/L	0.01
Spiked Amount	25.000	Range	70 - 120	Recovery	=	68.560%#
41) Toluene-d8	10.323	98	20733	15.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	61.480%
43) trans-1,3-Dichloroprop...	10.585	79	2996	15.917	ug/L	0.01
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.680%
47) 2-Hexanone-d5	10.927	63	1746	23.811	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	47.620%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	5842	17.236	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	68.960%
66) 1,2-Dichlorobenzene-d4	13.853	152	5943	19.194	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111122\
Data File : VW025040.D
Acq On : 11 Nov 2022 17:21
Operator : SY/VA
Sample : N5603-05
Misc : 5.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AH6

Quant Time: Nov 12 00:49:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Nov 12 00:48:53 2022
Response via : Initial Calibration

