

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111125\
 Data File : VW032433.D
 Acq On : 11 Nov 2025 16:05
 Operator : SY/MD
 Sample : Q3598-01
 Misc : 5.54g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-CONCRETE

Quant Time: Nov 12 03:51:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	269242	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.849	114	667437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	677738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	350516	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	316116	91.069	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	182.140%#
35) Dibromofluoromethane	7.904	113	265576	66.459	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	132.920%
50) Toluene-d8	10.325	98	1118070	72.932	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	145.860%#
62) 4-Bromofluorobenzene	12.617	95	501183	86.542	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	173.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111125\
Data File : VW032433.D
Acq On : 11 Nov 2025 16:05
Operator : SY/MD
Sample : Q3598-01
Misc : 5.54g/5mL/MSVOA_W/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-CONCRETE

Quant Time: Nov 12 03:51:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

