



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111125\
Data File : VW032414.D
Acq On : 11 Nov 2025 07:26
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Title : SW846 8260
Last Update : Tue Nov 11 03:48:21 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1748

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.1	60565	PASS
75	95	30	60	49.2	155768	PASS
95	95	100	100	100.0	316565	PASS
96	95	5	9	6.7	21304	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	66.0	208832	PASS
175	174	5	9	7.2	15131	PASS
176	174	95	101	98.5	205696	PASS
177	176	5	9	6.6	13673	PASS