

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111418\
 Data File : VW006885.D
 Acq On : 14 Nov 2018 23:52
 Operator : SY/AP
 Sample : J5944-03MSD
 Misc : 4.80G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ETHS2MSD

Quant Time: Nov 15 10:35:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111418S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 15 03:38:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	176459	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	170037	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	75984	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	38456	24.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.20%
7) Chloroethane-d5	2.89	69	29738	26.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.48%
10) 1,1-Dichloroethene-d2	4.03	63	99110	23.68	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	94.72%
20) 2-Butanone-d5	7.09	46	28494	33.00	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.00%
24) Chloroform-d	7.65	84	92623	23.71	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.84%
26) 1,2-Dichloroethane-d4	8.31	65	54856	24.88	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.52%
29) Benzene-d6	8.28	84	206620	23.25	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.00%
33) 1,2-Dichloropropane-d6	9.27	67	63593	23.37	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.48%
37) Toluene-d8	10.33	98	212640	23.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.44%
38) trans-1,3-Dichloropropene-	10.58	79	26203	22.42	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.68%
39) 2-Hexanone-d5	10.93	63	23219	34.25	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.50%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	48402	20.55	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.20%
61) 1,2-Dichlorobenzene-d4	13.86	152	68561	23.83	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.32%

Target Compounds					Ovalue
12) 1,1-Dichloroethene	4.04	96	48100	23.860 ug/L	96
16) Methylene chloride	4.92	84	6700	2.171 ug/L	96
34) Benzene	8.33	78	226123	23.779 ug/L	100
35) Trichloroethene	9.09	95	58838	22.717 ug/L	99
44) Toluene	10.39	91	269928	24.791 ug/L	99
52) Chlorobenzene	11.66	112	170344	24.874 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111418\
Data File : VW006885.D
Acq On : 14 Nov 2018 23:52
Operator : SY/AP
Sample : J5944-03MSD
Misc : 4.80G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETHS2MSD

Quant Time: Nov 15 10:35:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111418S.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 15 03:38:37 2018
Response via : Initial Calibration

