

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111422\
 Data File : VW025047.D
 Acq On : 14 Nov 2022 13:52
 Operator : SY/VA
 Sample : VIBLK583
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK583

Quant Time: Nov 14 23:45:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 14 23:45:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	296406	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	271181	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	128722	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	56355	16.939	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.760%
7) Chloroethane-d5	2.892	69	39787	20.415	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.640%
11) 1,1-Dichloroethene-d2	4.019	63	129502	15.626	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.520%
21) 2-Butanone-d5	7.074	46	42656	37.989	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.980%
24) Chloroform-d	7.647	84	202873	22.553	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	8.305	65	110565	21.958	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.840%
32) Benzene-d6	8.275	84	384474	22.120	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.480%
36) 1,2-Dichloropropane-d6	9.275	67	110311	21.660	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	86.640%
41) Toluene-d8	10.323	98	349392	21.591	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.360%
43) trans-1,3-Dichloroprop...	10.579	79	47079	20.847	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.400%
47) 2-Hexanone-d5	10.927	63	32929	37.429	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.860%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	79566	19.566	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	110630	22.364	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111422\
Data File : VW025047.D
Acq On : 14 Nov 2022 13:52
Operator : SY/VA
Sample : VIBLK583
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK583

Quant Time: Nov 14 23:45:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 14 23:45:04 2022
Response via : Initial Calibration

