

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111524\
 Data File : VW031029.D
 Acq On : 15 Nov 2024 16:58
 Operator : SY/MD
 Sample : P4872-03
 Misc : 5.22g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E29X3

Quant Time: Nov 16 00:38:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 16 00:35:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	568797	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	533570	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	240674	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	116005	12.435	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	49.760%
7) Chloroethane-d5	2.899	69	95789	13.388	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.560%
11) 1,1-Dichloroethene-d2	4.021	65	57956	15.056	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.240%
21) 2-Butanone-d5	7.075	46	77136	50.245	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.500%
24) Chloroform-d	7.648	84	300818	17.940	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	71.760%
26) 1,2-Dichloroethane-d4	8.301	65	158910	19.060	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.240%
32) Benzene-d6	8.270	84	578675	17.457	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.840%
36) 1,2-Dichloropropane-d6	9.270	67	180707	18.493	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	73.960%
41) Toluene-d8	10.319	98	518071	17.424	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.680%
43) trans-1,3-Dichloroprop...	10.575	79	70570	17.817	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	71.280%
47) 2-Hexanone-d5	10.922	63	64224	50.713	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.420%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	158327	23.256	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.040%
66) 1,2-Dichlorobenzene-d4	13.848	152	184582	21.055	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111524\
Data File : VW031029.D
Acq On : 15 Nov 2024 16:58
Operator : SY/MD
Sample : P4872-03
Misc : 5.22g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E29X3

Quant Time: Nov 16 00:38:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Nov 16 00:35:12 2024
Response via : Initial Calibration

