

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111622\
 Data File : VW025060.D
 Acq On : 16 Nov 2022 16:02
 Operator : SY/VA
 Sample : VIBLK585
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK585

Quant Time: Nov 17 00:40:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 17 00:39:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	464499	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	424315	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	203053	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	104173	19.981	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	79.920%		
7) Chloroethane-d5	2.885	69	75973	24.875	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	99.520%		
11) 1,1-Dichloroethene-d2	4.019	63	181831	14.001	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	56.000%		
21) 2-Butanone-d5	7.080	46	92356	52.486	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	104.980%		
24) Chloroform-d	7.653	84	312957	22.201	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	88.800%		
26) 1,2-Dichloroethane-d4	8.305	65	164336	20.826	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.320%		
32) Benzene-d6	8.275	84	629421	23.143	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	92.560%		
36) 1,2-Dichloropropane-d6	9.274	67	198033	24.851	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	99.400%		
41) Toluene-d8	10.323	98	555104	21.923	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	87.680%		
43) trans-1,3-Dichloroprop...	10.579	79	73911	20.917	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	83.680%		
47) 2-Hexanone-d5	10.920	63	71504	51.944	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	103.880%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	158935	24.979	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	99.920%		
66) 1,2-Dichlorobenzene-d4	13.853	152	182304	23.363	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	93.440%		
Target Compounds						Qvalue
16) Methylene chloride	4.922	84	35198	4.268	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111622\
Data File : VW025060.D
Acq On : 16 Nov 2022 16:02
Operator : SY/VA
Sample : VIBLK585
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK585

Quant Time: Nov 17 00:40:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 17 00:39:48 2022
Response via : Initial Calibration

