

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111725\
 Data File : VW032456.D
 Acq On : 17 Nov 2025 14:33
 Operator : SY/MD
 Sample : Q3633-01
 Misc : 5.14g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ETGI-369

Quant Time: Nov 18 03:04:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	192738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	459125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	492831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	234135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	164596	66.240	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	132.480%	
35) Dibromofluoromethane	7.904	113	147884	53.798	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	107.600%	
50) Toluene-d8	10.325	98	521048	49.409	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	98.820%	
62) 4-Bromofluorobenzene	12.617	95	209467	52.581	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	105.160%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.954	84	14456	5.437	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111725\
Data File : VW032456.D
Acq On : 17 Nov 2025 14:33
Operator : SY/MD
Sample : Q3633-01
Misc : 5.14g/5mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETGI-369

Quant Time: Nov 18 03:04:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

