

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
 Data File : VW025087.D
 Acq On : 18 Nov 2022 09:35
 Operator : SY/VA
 Sample : VW1118SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 19 04:31:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 17 23:58:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	439051	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	400839	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	187258	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	93186	18.909	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.640%
7) Chloroethane-d5	2.891	69	73548	25.477	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.920%
11) 1,1-Dichloroethene-d2	4.025	63	178291	14.524	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.080%
21) 2-Butanone-d5	7.080	46	71919	43.241	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.480%
24) Chloroform-d	7.653	84	289065	21.694	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.760%
26) 1,2-Dichloroethane-d4	8.311	65	148268	19.879	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.520%
32) Benzene-d6	8.275	84	582449	22.670	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.680%
36) 1,2-Dichloropropane-d6	9.274	67	180828	24.021	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.080%
41) Toluene-d8	10.323	98	511418	21.381	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.520%
43) trans-1,3-Dichloroprop...	10.579	79	63852	19.128	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.520%
47) 2-Hexanone-d5	10.926	63	55061	42.342	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.680%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	134952	22.452	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	163212	22.680	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
Data File : VW025087.D
Acq On : 18 Nov 2022 09:35
Operator : SY/VA
Sample : VW1118SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 19 04:31:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 17 23:58:11 2022
Response via : Initial Calibration

