

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
 Data File : VW025090.D
 Acq On : 18 Nov 2022 10:50
 Operator : SY/VA
 Sample : N5658-11
 Misc : 5.60g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 19 04:31:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 17 23:58:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	461925	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	381147	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	135686	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	89820	17.324	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.280%
7) Chloroethane-d5	2.885	69	70816	23.316	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.280%
11) 1,1-Dichloroethene-d2	4.019	63	172678	13.370	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.480%
21) 2-Butanone-d5	7.080	46	77193	44.114	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.220%
24) Chloroform-d	7.653	84	292483	20.864	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.440%
26) 1,2-Dichloroethane-d4	8.305	65	144265	18.384	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.520%
32) Benzene-d6	8.275	84	562645	23.031	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.120%
36) 1,2-Dichloropropane-d6	9.274	67	180332	25.192	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.760%
41) Toluene-d8	10.323	98	488046	21.458	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.840%
43) trans-1,3-Dichloroprop...	10.579	79	59131	18.629	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.520%
47) 2-Hexanone-d5	10.926	63	57453	46.464	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.920%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	128050	22.404	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	118101	22.649	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
Data File : VW025090.D
Acq On : 18 Nov 2022 10:50
Operator : SY/VA
Sample : N5658-11
Misc : 5.60g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 19 04:31:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 17 23:58:11 2022
Response via : Initial Calibration

