

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
 Data File : VW025091.D
 Acq On : 18 Nov 2022 11:14
 Operator : SY/VA
 Sample : N5658-09
 Misc : 4.40g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ESQS9

Quant Time: Nov 19 04:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 17 23:58:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	486928	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	374971	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	109743	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	102577	18.768	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.080%
7) Chloroethane-d5	2.885	69	76114	23.773	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.080%
11) 1,1-Dichloroethene-d2	4.019	63	196625	14.443	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.760%
21) 2-Butanone-d5	7.080	46	75474	40.917	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.840%
24) Chloroform-d	7.653	84	329768	22.316	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.280%
26) 1,2-Dichloroethane-d4	8.305	65	157724	19.067	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.280%
32) Benzene-d6	8.275	84	652046	27.130	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.520%
36) 1,2-Dichloropropane-d6	9.274	67	200390	28.456	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.840%
41) Toluene-d8	10.323	98	545906	24.397	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.600%
43) trans-1,3-Dichloroprop...	10.579	79	59816	19.156	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.640%
47) 2-Hexanone-d5	10.920	63	54033	44.417	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.840%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	125931	22.396	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	103088	24.444	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
Data File : VW025091.D
Acq On : 18 Nov 2022 11:14
Operator : SY/VA
Sample : N5658-09
Misc : 4.40g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ESQS9

Quant Time: Nov 19 04:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 17 23:58:11 2022
Response via : Initial Calibration

