

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
 Data File : VW025101.D
 Acq On : 18 Nov 2022 16:29
 Operator : SY/VA
 Sample : N5712-04RE
 Misc : 4.86g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGC77RE

Quant Time: Nov 19 04:33:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 17 23:58:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	212601	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	180417	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	58063	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	43556	18.253	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.000%
7) Chloroethane-d5	2.892	69	35407	25.329	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.320%
11) 1,1-Dichloroethene-d2	4.019	63	85038	14.306	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.240%
21) 2-Butanone-d5	7.080	46	33682	41.821	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.640%
24) Chloroform-d	7.653	84	137634	21.332	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.320%
26) 1,2-Dichloroethane-d4	8.305	65	69911	19.357	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.440%
32) Benzene-d6	8.275	84	264654	22.886	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.560%
36) 1,2-Dichloropropane-d6	9.275	67	83989	24.788	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.160%
41) Toluene-d8	10.323	98	218403	20.286	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.160%
43) trans-1,3-Dichloroprop...	10.579	79	27592	18.365	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.440%
47) 2-Hexanone-d5	10.927	63	23409	39.994	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.980%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	55478	20.506	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	44105	19.766	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111822\
Data File : VW025101.D
Acq On : 18 Nov 2022 16:29
Operator : SY/VA
Sample : N5712-04RE
Misc : 4.86g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGC77RE

Quant Time: Nov 19 04:33:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 17 23:58:11 2022
Response via : Initial Calibration

