

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111825\
 Data File : VW032476.D
 Acq On : 18 Nov 2025 13:18
 Operator : SY/MD
 Sample : Q3638-01
 Misc : 3.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 MOO-25-0327

Quant Time: Nov 19 00:26:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	217499	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	524737	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	560958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	282583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	162272	57.870	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	115.740%	
35) Dibromofluoromethane	7.898	113	150774	47.991	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	95.980%	
50) Toluene-d8	10.325	98	528531	43.852	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	87.700%	
62) 4-Bromofluorobenzene	12.617	95	236603	51.966	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	103.940%	
Target Compounds						
16) Acetone	4.167	43	15949	19.012	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111825\
Data File : VW032476.D
Acq On : 18 Nov 2025 13:18
Operator : SY/MD
Sample : Q3638-01
Misc : 3.00g/5mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
MOO-25-0327

Quant Time: Nov 19 00:26:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

