

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
 Data File : VW032491.D
 Acq On : 19 Nov 2025 10:36
 Operator : SY/MD
 Sample : VW1119SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1119SBL01

Quant Time: Nov 20 01:15:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	218109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	508852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	541801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	260882	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	170243	60.543	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	121.080%
35) Dibromofluoromethane	7.898	113	157341	51.644	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	103.280%
50) Toluene-d8	10.325	98	531813	45.501	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	91.000%
62) 4-Bromofluorobenzene	12.617	95	240405	54.449	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	108.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
Data File : VW032491.D
Acq On : 19 Nov 2025 10:36
Operator : SY/MD
Sample : VW1119SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1119SBL01

Quant Time: Nov 20 01:15:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

