

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
 Data File : VW032497.D
 Acq On : 19 Nov 2025 13:14
 Operator : SY/MD
 Sample : Q3649-04
 Misc : 6.21g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DUPE-0.5-1.0-20251114

Quant Time: Nov 20 01:18:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	197627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	485828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	516048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	249401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	153222	60.137	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	120.280%
35) Dibromofluoromethane	7.904	113	141253	48.561	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	97.120%
50) Toluene-d8	10.325	98	494342	44.300	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	88.600%
62) 4-Bromofluorobenzene	12.617	95	216521	51.364	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	102.720%
Target Compounds						
16) Acetone	4.155	43	24175	31.715	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
Data File : VW032497.D
Acq On : 19 Nov 2025 13:14
Operator : SY/MD
Sample : Q3649-04
Misc : 6.21g/5mL/MSVOA_W/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DUPE-0.5-1.0-20251114

Quant Time: Nov 20 01:18:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

