

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
 Data File : VW032509.D
 Acq On : 19 Nov 2025 17:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 01:21:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.254	0.256	-0.8	74	0.00
3 P	Chloromethane	0.435	0.452	-3.9	79	0.00
4 C	Vinyl Chloride	0.549	0.540	1.6#	72	0.00
5 T	Bromomethane	0.367	0.370	-0.8	74	0.00
6 T	Chloroethane	0.359	0.367	-2.2	74	0.00
7 T	Trichlorofluoromethane	0.430	0.406	5.6	63	0.00
8 T	Diethyl Ether	0.352	0.370	-5.1	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.469	0.471	-0.4	72	0.00
10 T	Methyl Iodide	0.755	0.767	-1.6	74	0.00
11 T	Tert butyl alcohol	0.056	0.051	8.9	68	0.00
12 CM	1,1-Dichloroethene	0.537	0.540	-0.6#	72	0.00
13 T	Acrolein	0.094	0.053	43.6#	43#	0.00
14 T	Allyl chloride	1.034	1.021	1.3	70	0.00
15 T	Acrylonitrile	0.203	0.217	-6.9	77	0.00
16 T	Acetone	0.193	0.180	6.7	69	0.00
17 T	Carbon Disulfide	1.588	1.539	3.1	69	0.00
18 T	Methyl Acetate	0.478	0.668	-39.7#	103	0.00
19 T	Methyl tert-butyl Ether	1.133	1.137	-0.4	72	0.00
20 T	Methylene Chloride	0.690	0.718	-4.1	81	0.00
21 T	trans-1,2-Dichloroethene	0.615	0.644	-4.7	77	0.00
22 T	Diisopropyl ether	2.104	2.264	-7.6	78	0.00
23 T	Vinyl Acetate	1.442	1.414	1.9	69	0.00
24 P	1,1-Dichloroethane	1.153	1.259	-9.2	79	0.00
25 T	2-Butanone	0.284	0.286	-0.7	73	0.00
26 T	2,2-Dichloropropane	0.643	0.582	9.5	67	0.00
27 T	cis-1,2-Dichloroethene	0.734	0.787	-7.2	77	0.00
28 T	Bromochloromethane	0.545	0.602	-10.5	86	0.00
29 T	Tetrahydrofuran	0.180	0.195	-8.3	77	0.00
30 C	Chloroform	1.129	1.268	-12.3#	82	0.00
31 T	Cyclohexane	1.066	1.023	4.0	71	0.00
32 T	1,1,1-Trichloroethane	0.806	0.841	-4.3	74	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.662	-2.6	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.299	0.311	-4.0	80	0.00
36 T	1,1-Dichloropropene	0.441	0.474	-7.5	77	0.00
37 T	Ethyl Acetate	0.328	0.341	-4.0	75	0.00
38 T	Carbon Tetrachloride	0.392	0.415	-5.9	77	0.00
39 T	Methylcyclohexane	0.571	0.570	0.2	71	0.00
40 TM	Benzene	1.420	1.538	-8.3	79	0.00
41 T	Methacrylonitrile	0.180	0.185	-2.8	73	0.00
42 TM	1,2-Dichloroethane	0.417	0.471	-12.9	83	0.00
43 T	Isopropyl Acetate	0.581	0.600	-3.3	74	0.00
44 TM	Trichloroethene	0.331	0.347	-4.8	77	0.00
45 C	1,2-Dichloropropane	0.354	0.393	-11.0#	82	0.00
46 T	Dibromomethane	0.207	0.225	-8.7	79	0.00
47 T	Bromodichloromethane	0.471	0.517	-9.8	79	0.00
48 T	Methyl methacrylate	0.276	0.286	-3.6	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
 Data File : VW032509.D
 Acq On : 19 Nov 2025 17:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 01:21:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	75	0.00
50 S	Toluene-d8	1.148	1.191	-3.7	78	0.00
51 T	4-Methyl-2-Pentanone	0.321	0.352	-9.7	78	0.00
52 CM	Toluene	0.877	0.949	-8.2#	79	0.00
53 T	t-1,3-Dichloropropene	0.486	0.502	-3.3	74	0.00
54 T	cis-1,3-Dichloropropene	0.562	0.590	-5.0	75	0.00
55 T	1,1,2-Trichloroethane	0.284	0.309	-8.8	78	0.00
56 T	Ethyl methacrylate	0.435	0.467	-7.4	75	0.00
57 T	1,3-Dichloropropane	0.506	0.553	-9.3	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.232	0.263	-13.4	86	0.00
59 T	2-Hexanone	0.235	0.246	-4.7	75	0.00
60 T	Dibromochloromethane	0.317	0.339	-6.9	78	0.00
61 T	1,2-Dibromoethane	0.276	0.296	-7.2	78	0.00
62 S	4-Bromofluorobenzene	0.434	0.443	-2.1	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.296	0.343	-15.9	88	0.00
65 PM	Chlorobenzene	1.078	1.157	-7.3	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.334	0.362	-8.4	82	0.00
67 C	Ethyl Benzene	1.865	1.949	-4.5#	77	0.00
68 T	m/p-Xylenes	0.700	0.747	-6.7	79	0.00
69 T	o-Xylene	0.670	0.708	-5.7	80	0.00
70 T	Styrene	1.153	1.261	-9.4	81	0.00
71 P	Bromoform	0.198	0.194	2.0	74	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.672	3.691	-0.5	77	0.00
74 T	N-amyl acetate	1.276	1.239	2.9	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.871	0.908	-4.2	81	0.00
76 T	1,2,3-Trichloropropane	0.686	0.651	5.1	80	0.00
77 T	Bromobenzene	0.869	0.858	1.3	75	0.00
78 T	n-propylbenzene	4.472	4.639	-3.7	80	0.00
79 T	2-Chlorotoluene	2.730	2.792	-2.3	80	0.00
80 T	1,3,5-Trimethylbenzene	3.030	3.089	-1.9	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.299	0.288	3.7	73	0.00
82 T	4-Chlorotoluene	2.839	2.882	-1.5	79	0.00
83 T	tert-Butylbenzene	2.635	2.587	1.8	74	0.00
84 T	1,2,4-Trimethylbenzene	3.081	3.200	-3.9	81	0.00
85 T	sec-Butylbenzene	3.925	3.853	1.8	76	0.00
86 T	p-Isopropyltoluene	3.241	3.224	0.5	78	0.00
87 T	1,3-Dichlorobenzene	1.648	1.705	-3.5	82	0.00
88 T	1,4-Dichlorobenzene	1.689	1.677	0.7	78	0.00
89 T	n-Butylbenzene	3.166	3.116	1.6	76	0.00
90 T	Hexachloroethane	0.584	0.581	0.5	77	0.00
91 T	1,2-Dichlorobenzene	1.513	1.538	-1.7	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.150	0.146	2.7	76	0.00
93 T	1,2,4-Trichlorobenzene	0.963	0.911	5.4	73	0.00
94 T	Hexachlorobutadiene	0.400	0.340	15.0	67	0.00
95 T	Naphthalene	2.401	2.313	3.7	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111925\
Data File : VW032509.D
Acq On : 19 Nov 2025 17:37
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 01:21:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.874	0.852	2.5	78	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6