

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112025\
 Data File : VW032511.D
 Acq On : 20 Nov 2025 09:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 01:14:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.254	0.246	3.1	71	0.00
3 P	Chloromethane	0.435	0.464	-6.7	81	0.00
4 C	Vinyl Chloride	0.549	0.632	-15.1#	84	0.00
5 T	Bromomethane	0.367	0.418	-13.9	83	0.00
6 T	Chloroethane	0.359	0.407	-13.4	82	0.00
7 T	Trichlorofluoromethane	0.430	0.461	-7.2	71	0.00
8 T	Diethyl Ether	0.352	0.373	-6.0	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.469	0.542	-15.6	83	0.00
10 T	Methyl Iodide	0.755	0.825	-9.3	79	0.00
11 T	Tert butyl alcohol	0.056	0.050	10.7	66	-0.01
12 CM	1,1-Dichloroethene	0.537	0.601	-11.9#	79	0.00
13 T	Acrolein	0.094	0.060	36.2#	47#	0.00
14 T	Allyl chloride	1.034	1.131	-9.4	77	0.00
15 T	Acrylonitrile	0.203	0.220	-8.4	77	-0.01
16 T	Acetone	0.193	0.218	-13.0	83	0.00
17 T	Carbon Disulfide	1.588	1.764	-11.1	78	0.00
18 T	Methyl Acetate	0.478	0.529	-10.7	81	0.00
19 T	Methyl tert-butyl Ether	1.133	1.204	-6.3	75	0.00
20 T	Methylene Chloride	0.690	0.762	-10.4	85	0.00
21 T	trans-1,2-Dichloroethene	0.615	0.689	-12.0	81	0.00
22 T	Diisopropyl ether	2.104	2.342	-11.3	80	0.00
23 T	Vinyl Acetate	1.442	1.550	-7.5	76	0.00
24 P	1,1-Dichloroethane	1.153	1.327	-15.1	82	0.00
25 T	2-Butanone	0.284	0.295	-3.9	75	0.00
26 T	2,2-Dichloropropane	0.643	0.694	-7.9	80	0.00
27 T	cis-1,2-Dichloroethene	0.734	0.812	-10.6	79	0.00
28 T	Bromochloromethane	0.545	0.605	-11.0	86	0.00
29 T	Tetrahydrofuran	0.180	0.194	-7.8	76	0.00
30 C	Chloroform	1.129	1.318	-16.7#	85	0.00
31 T	Cyclohexane	1.066	1.116	-4.7	77	0.00
32 T	1,1,1-Trichloroethane	0.806	0.911	-13.0	79	0.00
33 S	1,2-Dichloroethane-d4	0.645	0.681	-5.6	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.299	0.319	-6.7	81	0.00
36 T	1,1-Dichloropropene	0.441	0.500	-13.4	79	0.00
37 T	Ethyl Acetate	0.328	0.360	-9.8	77	0.00
38 T	Carbon Tetrachloride	0.392	0.459	-17.1	83	0.00
39 T	Methylcyclohexane	0.571	0.635	-11.2	77	0.00
40 TM	Benzene	1.420	1.639	-15.4	83	0.00
41 T	Methacrylonitrile	0.180	0.185	-2.8	72	0.00
42 TM	1,2-Dichloroethane	0.417	0.489	-17.3	84	0.00
43 T	Isopropyl Acetate	0.581	0.617	-6.2	74	0.00
44 TM	Trichloroethene	0.331	0.363	-9.7	79	0.00
45 C	1,2-Dichloropropane	0.354	0.409	-15.5#	83	0.00
46 T	Dibromomethane	0.207	0.240	-15.9	82	0.00
47 T	Bromodichloromethane	0.471	0.540	-14.6	81	0.00
48 T	Methyl methacrylate	0.276	0.294	-6.5	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112025\
 Data File : VW032511.D
 Acq On : 20 Nov 2025 09:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 01:14:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	77	0.00
50 S	Toluene-d8	1.148	1.247	-8.6	80	0.00
51 T	4-Methyl-2-Pentanone	0.321	0.359	-11.8	77	0.00
52 CM	Toluene	0.877	1.011	-15.3#	82	0.00
53 T	t-1,3-Dichloropropene	0.486	0.534	-9.9	77	0.00
54 T	cis-1,3-Dichloropropene	0.562	0.643	-14.4	80	0.00
55 T	1,1,2-Trichloroethane	0.284	0.322	-13.4	80	0.00
56 T	Ethyl methacrylate	0.435	0.466	-7.1	74	0.00
57 T	1,3-Dichloropropane	0.506	0.566	-11.9	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.232	0.256	-10.3	82	0.00
59 T	2-Hexanone	0.235	0.257	-9.4	76	0.00
60 T	Dibromochloromethane	0.317	0.360	-13.6	82	0.00
61 T	1,2-Dibromoethane	0.276	0.308	-11.6	79	0.00
62 S	4-Bromofluorobenzene	0.434	0.448	-3.2	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.296	0.329	-11.1	83	0.00
65 PM	Chlorobenzene	1.078	1.187	-10.1	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.334	0.369	-10.5	82	0.00
67 C	Ethyl Benzene	1.865	2.096	-12.4#	81	0.00
68 T	m/p-Xylenes	0.700	0.782	-11.7	81	0.00
69 T	o-Xylene	0.670	0.743	-10.9	82	0.00
70 T	Styrene	1.153	1.322	-14.7	83	0.00
71 P	Bromoform	0.198	0.210	-6.1	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.672	3.879	-5.6	80	0.00
74 T	N-amyl acetate	1.276	1.270	0.5	74	0.00
75 P	1,1,2,2-Tetrachloroethane	0.871	0.904	-3.8	80	0.00
76 T	1,2,3-Trichloropropane	0.686	0.642	6.4	78	0.00
77 T	Bromobenzene	0.869	0.910	-4.7	79	0.00
78 T	n-propylbenzene	4.472	4.897	-9.5	83	0.00
79 T	2-Chlorotoluene	2.730	2.944	-7.8	84	0.00
80 T	1,3,5-Trimethylbenzene	3.030	3.220	-6.3	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.299	0.306	-2.3	76	0.00
82 T	4-Chlorotoluene	2.839	3.011	-6.1	82	0.00
83 T	tert-Butylbenzene	2.635	2.827	-7.3	80	0.00
84 T	1,2,4-Trimethylbenzene	3.081	3.302	-7.2	82	0.00
85 T	sec-Butylbenzene	3.925	4.145	-5.6	81	0.00
86 T	p-Isopropyltoluene	3.241	3.432	-5.9	82	0.00
87 T	1,3-Dichlorobenzene	1.648	1.741	-5.6	83	0.00
88 T	1,4-Dichlorobenzene	1.689	1.776	-5.2	82	0.00
89 T	n-Butylbenzene	3.166	3.382	-6.8	82	0.00
90 T	Hexachloroethane	0.584	0.597	-2.2	78	0.00
91 T	1,2-Dichlorobenzene	1.513	1.611	-6.5	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.150	0.148	1.3	76	0.00
93 T	1,2,4-Trichlorobenzene	0.963	0.933	3.1	74	0.00
94 T	Hexachlorobutadiene	0.400	0.375	6.3	73	0.00
95 T	Naphthalene	2.401	2.319	3.4	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112025\
Data File : VW032511.D
Acq On : 20 Nov 2025 09:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Nov 21 01:14:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.874	0.879	-0.6	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6